

Certificate No.

24PTP\_0104240192

Page 1 of 3

### **Certificate of Calibration**

EQUIPMENT : MONITORS, BEDSIDE, PHYSIOLOGIC

ID CODE : PTP00184

MANUFACTURER : PHILIPS

MODEL : INTELLIVUE MP20(M8001A)

SERIAL No. : DE72876221

LOCATION : Patient Wards 17

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

CALIBRATED DATE : 18 APRIL 2024

ISSUE DATE : 3 MAY 2024

Usanee

Approved by :

USANEE CHALONGPHOKSILCHAI

PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

### **CONDITION OF THIS RESULT OF TEST**

Performed by:

#### 1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEVitalsign SimulatorFlukeProSim8254103123MD7777 Jun 2024

#### 2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

## 3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



*Certificate No.* 24PTP\_0104240192

Page 2 of 3

## **Calibration Report**

EQUIPMENT : MONITORS, BEDSIDE, PHYSIOLOGIC

ID CODE : PTP00184

MANUFACTURER : PHILIPS

MODEL : INTELLIVUE MP20(M8001A)

SERIAL No. : DE72876221

DATE OF CALIBRATION : 18 APRIL 2024

ENVIRONMENT : TEMPERATURE 25 °C

RELATIVE HUMIDITY 55 %

### **PROCEDURE USED:**

This instrument was calibration by comparison with standard

### **MEASURMENT RESULT:**

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

Systolic(NIBP)						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
60	62.00	2.00	3.33	± 0.9238		
120	120.00	0.00	0.00	± 1.2702		
200	200.00	0.00	0.00	± 1.7321		

Diastolic(NIBP)						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
30	33.00	3.00	10.00	± 0.7506		
80	84.00	4.00	5.00	± 1.0393		
150	151.00	1.00	0.67	± 1.4434		

Heart Rate						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(BPM)	(BPM)	(BPM)	(%)	(BPM)		
60	60.00	0.00	0.00	± 0.0150		
80	80.00	0.00	0.00	± 0.0460		
120	120.00	0.00	0.00	± 0.0460		

FI-BME-NHS-CP-029/1 Rev.03 Page 2/3 Issued Date 20/07/2023



*Certificate No.* 24PTP\_0104240192

Page 3 of 3

# **Calibration Report**

EQUIPMENT : MONITORS, BEDSIDE, PHYSIOLOGIC

ID CODE : PTP00184 SERIAL No. : DE72876221

%Spo2						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(%SPO <sub>2</sub> )	(%SPO <sub>2</sub> )	(%SPO <sub>2</sub> )	(%)	(%SPO <sub>2</sub> )		
90	90.00	0.00	0.00	± 0.1039		
96	96.00	0.00	0.00	± 0.1109		
100	100.00	0.00	0.00	± 0.1155		

Temperature						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(°C)	(°C)	(°C)	(%)	(°C)		
35	35.00	0.00	0.00	± 0.0051		
37	37.00	0.00	0.00	± 0.0051		
39	39.00	0.00	0.00	± 0.0055		

Systolic(IBP)						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
10	10.00	0.00	0.00	± 0.1155		
24	24.00	0.00	0.00	± 0.2771		
120	119.00	-1.00	-0.83	± 1.3857		

Diastolic(IBP)						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
2	2.00	0.00	0.00	± 0.0231		
10	10.00	0.00	0.00	± 0.1155		
80	80.00	0.00	0.00	± 0.9238		

### **UUC\***: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%