

Report No.

24PTP\_01PM04240187

Page 1 of 2

## **Preventive Maintenance Action Report**

EQUIPMENT : MODULE, ECG/SPO2/NIBP

ID CODE : PTP00316

MANUFACTURER : PHILIPS

MODEL : M3001A

SERIAL No. : DE9070A3GJ

LOCATION : I.C.U

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

PM. DATE : 18 APRIL 2024

USANEE CHALONGPHOKSILCHAI

ISSUE DATE : 2 MAY 2024

Performed by:

Approved by :

PHUWASIN YIWKIM

## **CONDITION OF THIS RESULT OF TEST**

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE

2. THIS CERTIFICATION IS TRACEABLE TO:

-

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Report No. 24PTP\_01PM04240187

Page 2 of 2

## **Preventive Maintenance Action Report**

EQUIPMENT : MODULE, ECG/SPO2/NIBP

ID CODE : PTP00316

SERIAL No. : DE9070A3GJ

## 1.Report

| No. | List                               | Pass | Fail | None | Comment |
|-----|------------------------------------|------|------|------|---------|
| 1   | This Module Calibrate With Monitor | /    |      |      |         |

**2.Remark** MODULE OF ID Code :PTP02717

 $Electrical\ Safety\ Test: Master\ used\ Electrical\ Safety\ Analyzer\ (\ Metron\ /\ QA-90\ )\ according\ to\ IEC.60601.1$ 

FI-BME-NHS-CP-001/3 Rev.09 Page 2/2 Issued Date 20/07/2023