

Report No. 24PTP_01PM04240264

Page 1 of 2

Preventive Maintenance Action Report

EQUIPMENT : WARMING UNITS, PATIENT, FORCED-AIR

ID CODE : PTP00771

MANUFACTURER : BAIR HUGGER

MODEL : 775

SERIAL No. : 85703

LOCATION : Anesthesia

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

PM. DATE : 24 APRIL 2024

ISSUE DATE : 1 MAY 2024

Supatcha

Performed by : _____ Approved by : _____

SUPATCHA BUAPUEAN PHUWASIN YIWKIM

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEElectrical SafetyFlukeESA615349403423MD7747 Jun 2024

Analyzer

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Report No. 24PTP_01PM04240264

Page 2 of 2

Preventive Maintenance Action Report

EQUIPMENT : WARMING UNITS, PATIENT, FORCED-AIR

ID CODE : PTP00771

SERIAL No. : 85703

1. Qualitative Tests

No.	List	Pass	Fail	None	Comment
1	Chassis/Housing	/			
2	Mount	/			
3	Casters/Brakes	/			
4	AC Plug	/			
5	Line Cord	/			
6	Strain Reliefs	/			
7	Probes	/			
8	Controls/Switches	/			
9	Heater	/			
10	Blower	/			
11	Indicators/Displays	/			
12	Alarms	/			
13	Audible Signals	/			
14	Labeling	/			
15	Blankets	/			

2. Quantitative Tests

No.	List	Pass	Fail	None	Comment
1	Grounding Resistance ≤0.5 Ω	/			0.361 Ω
2	Chassis Leakage Current ≤500 μA	/			69.7 μΑ
3	High-Temperature Protection 53°C ±3°C (127°F	/			
	±5°F)				
4	Temperature Accuracy ±1°C (5°F) displayed value	/			

3. Preventive Maintenance

No.	List	Pass	Fail	None	Comment
1	Clean the exterior and accessories	/			
2	Lubricate			/	
3	Adjust			/	
4	Replace			/	

Electrical Safety Test: Master used Electrical Safety Analyzer (Metron / QA-90) according to IEC.60601.1