

Certificate No. 24PTP_0104240191

Page 1 of 3

Certificate of Calibration

EQUIPMENT : MODULE, ECG/SPO2/NIBP

ID CODE : PTP00404

MANUFACTURER : PHILIPS

MODEL : M3001A

SERIAL No. : DE9070C8GG

LOCATION : I.C.U

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

CALIBRATED DATE : 18 APRIL 2024

ISSUE DATE : 2 MAY 2024

Performed by: _____ Approved by: ____

USANEE CHALONGPHOKSILCHAI PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEVitalsign SimulatorFlukeProSim8254103123MD7777 Jun 2024

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PTP_0104240191

Page 2 of 3

Calibration Report

EQUIPMENT : MODULE, ECG/SPO2/NIBP

ID CODE : PTP00404

MANUFACTURER : PHILIPS

MODEL : M3001A

SERIAL No. : DE9070C8GG

DATE OF CALIBRATION : 18 APRIL 2024

ENVIRONMENT : TEMPERATURE 25 °C

RELATIVE HUMIDITY 55 %

PROCEDURE USED:

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

Systolic(NIBP)						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
60	63.00	3.00	5.00	± 0.9238		
120	114.00	-6.00	-5.00	± 1.2702		
200	198.00	-2.00	-1.00	± 1.7321		

Diastolic(NIBP)						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
30	32.00	2.00	6.67	± 0.7506		
80	81.00	1.00	1.25	± 1.0393		
150	150.00	0.00	0.00	± 1.4434		

Heart Rate						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(BPM)	(BPM)	(BPM)	(%)	(BPM)		
60	60.00	0.00	0.00	± 0.0150		
80	80.00	0.00	0.00	± 0.0460		
120	120.00	0.00	0.00	± 0.0460		

FI-BME-NHS-CP-029/1 Rev.03 Page 2/3 Issued Date 20/07/2023



Certificate No. 24PTP_0104240191

Page 3 of 3

Calibration Report

EQUIPMENT : MODULE, ECG/SPO2/NIBP

ID CODE : PTP00404

SERIAL No. : DE9070C8GG

%Spo2						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(%SPO ₂)	(%SPO ₂)	(%SPO ₂)	(%)	(%SPO ₂)		
90	90.00	0.00	0.00	± 0.1039		
96	96.00	0.00	0.00	± 0.1109		
100	100.00	0.00	0.00	± 0.1155		

Temperature						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(°C)	(°C)	(°C)	(%)	(°C)		
35	35.00	0.00	0.00	± 0.0051		
37	37.00	0.00	0.00	± 0.0051		
39	39.00	0.00	0.00	± 0.0055		

Systolic(IBP)						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
10	10.00	0.00	0.00	± 0.1155		
24	24.00	0.00	0.00	± 0.2771		
120	119.00	-1.00	-0.83	± 1.3857		

Diastolic(IBP)						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
2	2.00	0.00	0.00	± 0.0231		
10	10.00	0.00	0.00	± 0.1155		
80	80.00	0.00	0.00	± 0.9238		

UUC*: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-029/1 Rev.03 Page 3/3 Issued Date 20/07/2023