

Certificate No.

24PYT3\_DEC12240085

Page 1 of 3

### **Certificate of Calibration**

EQUIPMENT : MONITORS, PHYSIOLOGIC, VITAL SIGNS

ID CODE : PYT3\_05565

MANUFACTURER : PHILIPS

MODEL : SURESIGNS VS4

SERIAL No. : CN42749557

LOCATION : OPD COUNTER NURSE (CN คัดกรอง)

SUBMITTED BY : PHYATHAI 3 HOSPITAL

111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160

Tel: (662) 467-1111 Fax: (622) 467-1111

CALIBRATED DATE : 1 DECEMBER 2024
ISSUE DATE : 30 DECEMBER 2024

Performed by : Approved by : PUKARIN TONGKLIANG

SOM KONKAEW PUKARIN TONGKLIANG

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

### **CONDITION OF THIS RESULT OF TEST**

#### 1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATETester, Vital SignFLUKEPROSIM 8570205924MD54530 Apr 2025

Simulator

### 2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

### 3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PYT3\_DEC12240085

Page 2 of 3

## **Calibration Report**

EQUIPMENT : MONITORS, PHYSIOLOGIC, VITAL SIGNS

ID CODE : PYT3\_05565

MANUFACTURER : PHILIPS

MODEL : SURESIGNS VS4

SERIAL No. : CN42749557

DATE OF CALIBRATION : 1 DECEMBER 2024

ENVIRONMENT : TEMPERATURE 25 °C

RELATIVE HUMIDITY 55 %

### **PROCEDURE USED:**

This instrument was calibration by comparison with standard

### **MEASURMENT RESULT:**

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

Systolic						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
60	63.01	3.01	5.02	± 0.0619		
120	118.00	-2.00	-1.67	± 0.0950		
200	199.98	-0.02	-0.01	± 0.0837		

Diastolic						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
30	34.01	4.01	13.36	± 0.0158		
80	77.98	-2.02	-2.53	± 0.0639		
150	150.99	0.99	0.66	± 0.0625		

Heart Rate						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(BPM)	(BPM)	(BPM)	(%)	(BPM)		
60	59.99	-0.01	-0.02	± 0.0548		
80	80.01	0.01	0.02	± 0.1174		
120	119.99	-0.01	-0.01	± 0.1207		

FI-BME-NHS-CP-007/1 Rev.04 Page 2/3 Issued Date 20/07/2024



Certificate No. 24PYT3\_DEC12240085

Page 3 of 3

# **Calibration Report**

EQUIPMENT : MONITORS, PHYSIOLOGIC, VITAL SIGNS

ID CODE : PYT3\_05565

SERIAL No. : CN42749557

%Spo2						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(%SPO <sub>2</sub> )	(%SPO <sub>2</sub> )	(%SPO <sub>2</sub> )	(%)	(%SPO <sub>2</sub> )		
90	90.01	0.01	0.01	± 0.0358		
96	97.04	1.04	1.08	± 0.0435		
100	100.03	0.03	0.03	± 0.0776		

#### **UUC\***: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-007/1 Rev.04 Page 3/3 Issued Date 20/07/2024