

Report No.

24PYT3_DECPM12240133

Page 1 of 2

Preventive Maintenance Action Report

	. 5022020	
ISSUE DATE	: 30 DECEM	BER 2024
PM. DATE	: 4 DECEMB	ER 2024
	Tel: (662) 4	167-1111 Fax : (622) 467-1111
	111 Phet Kaser	n Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 1016
SUBMITTED BY	: РНҮАТНА	I 3 HOSPITAL
LOCATION	: AMBULAN	ICE (รถพยาบาล)
SERIAL No.	: -	
MODEL	: VAC-420	
MANUFACTURER	: FOCUS	
ID CODE	: PYT3_0059	4
EQUIPMENT	: ASPIRATO	RS, EMERGENCY

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE

2. THIS CERTIFICATION IS TRACEABLE TO:

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Report No. 24PYT3_DECPM12240133

Page 2 of 2

Preventive Maintenance Action Report

EQUIPMENT : ASPIRATORS, EMERGENCY

ID CODE : PYT3_00594

SERIAL No. : -

1. Qualitative Tests

No.	List	Pass	Fail	None	Comment
1	1 Chassis/Housing				
2	2 Tubes/Hoses				
3	3 Fittings/Connectors				
4	Filters			/	
5	5 Controls/Switches				
6	6 Overflow Protection			/	
7	7 Intermittent Operation			/	
8	Indicators/Display	/			
9	Labeling	/			
10	Accessories	/			

2. Quantitative Tests

No.	List	Pass	Fail	None	Comment
1	Maximum Free Flow surgical/tracheal ≥25 L/min,			/	
	thoracic ≥20 L/min				
2	Maximum Vacuum Low-volume>100 mmHg;			/	
	thoracic >50 mmHg ; surgical/tracheal ≥400 mmHg				
3	Vacuum Gauge Accuracy ±5% FSO(Full Scale	/			
	output)				

3. Preventive Maintenance

No.	List	Pass	Fail	None	Comment
1	Clean the exterior and accessories	/			
2	Lubricate			/	
3	Adjust			/	
4	Replace			/	