

Report No. 25BNH_PMPM01250238

Page 1 of 2

Preventive Maintenance Action Report

EQUIPMENT : ASPIRATORS, EMERGENCY

ID CODE : BNH_ME 02266

MANUFACTURER : CA-MI

MODEL : ASKIR 230/12VBR

SERIAL No. : 14963

LOCATION : Haemodialysis Department

SUBMITTED BY : BNH HOSPITAL

9/1, Convent Road, Silom Bangkok 10500

Tel:- Fax:-

PM. DATE : 28 JANUARY 2025

ISSUE DATE : 24 FEBRUARY 2025

Performed by: ______ Approved by:

SUPPAWIT KULDILOK PHUWASIN YIWKIM

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE ESA615 2856013 24MD2088 26 Dec 2025 **Electrical Safety** Fluke Analyzer Pressure Calibrator Fluke 717 30G 6087518 24P4270 24 Dec 2025

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)
- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Report No. 25BNH_PMPM01250238

Page 2 of 2

Preventive Maintenance Action Report

EQUIPMENT : ASPIRATORS, EMERGENCY

ID CODE : BNH_ME 02266

SERIAL No. : 14963

1. Qualitative Tests

No.	List	Pass	Fail	Comment
1	Chassis/Housing	/		
2	AC Plug	/		
3	Line Cord	/		
4	Strain Reliefs	/		
5	Tubes/Hoses	/		
6	Fittings/Connectors	/		
7	Filters	/		
8	Controls/Switches	/		
9	Overflow Protection	/		
10	Motor/Pump	/		
11	Battery/Charger	/		
12	Indicators/Display	/		
13	Labeling	/		
14	Accessories	/		

2. Quantitative Tests

No.	List	Pass	Fail	Comment
1	Chassis Leakage Current ≤500 µA	/		35.2 μΑ

3. Preventive Maintenance

No.	List	Pass	Fail	Comment
1	Clean the exterior and accessories	/		
2	Replace	/		Filter

4. Remark ผลการทดสอบค่า Vaccum (max 100 kPa) 50 kPa = 49.31 kPa