

Report No. 24PTP_01PM04240143

Page 1 of 3

Preventive Maintenance Action Report

EQUIPMENT : OXYGEN HIGH FLOW

ID CODE : PTP03799

MANUFACTURER : FISHER & PAYKEL HEALTHCARE ?

MODEL : AIRVO 2

SERIAL No. : 210417257936

LOCATION : I.C.U

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

PM. DATE : 18 APRIL 2024 ISSUE DATE : 1 MAY 2024

Chakkaphan

CHAKKAPHAN KHAMPHUMI

Approved by :

PHUWASIN YIWKIM

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEElectrical SafetyFlukeESA615349403423MD7747 Jun 2024

Analyzer

Performed by:

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Report No. 24PTP_01PM04240143

Page 2 of 3

Preventive Maintenance Action Report

EQUIPMENT : OXYGEN HIGH FLOW

ID CODE : PTP03799

SERIAL No. : 210417257936

1. Qualitative Tests

No.	List	Pass	Fail	None	Comment
1	Chassis/Housing	/			
2	Casters/Wheels	/			
3	AC Plug	/			
4	Line Cord	/			
5	Strain Reliefs	/			
6	Circuit Breaker/Fuse	/			
7	Tubes/Hoses	/			
8	Filters	/			
9	Controls/Switches	/			
10	Fan/Compressor	/			
11	Indicators/Displays	/			
12	Flowmeter	/			
13	Alarms	/			
14	Operational Sound	/			
15	Labeling	/			
16	Accessories	/			

2. Quantitative Tests

No.	List	Pass	Fail	None	Comment
1	Grounding Resistance ≤0.5Ω			/	
2	Chassis Leakage Current ≤500 μA	/			13.8 μΑ
3	Oxygen Concentration ≥90%	/			
4	Flow Accuracy ±10%	/			
5	Cycle Time mfr spec	/			

3. Preventive Maintenance

No.	List	Pass	Fail	None	Comment
1	Clean the exterior and accessories	/			
2	Lubricate			/	
3	Adjust			/	
4	Replace			/	



Report No. 24PTP_01PM04240143

Page 3 of 3

Preventive Maintenance Action Report

EQUIPMENT : OXYGEN HIGH FLOW

ID CODE : PTP03799

SERIAL No. : 210417257936

Electrical Safety Test: Master used Electrical Safety Analyzer (Metron / QA-90) according to IEC.60601.1

FI-BME-NHS-CP-068/3 Rev.03 Page 3/3 Issued Date 20/07/2023