

Certificate No.

24PTP_0104240217

Page 1 of 3

Certificate of Calibration

EQUIPMENT : MONITORS, PHYSIOLOGIC, VITAL SIGNS

ID CODE : PTP04523

MANUFACTURER : WELCH ALLYN

MODEL : CONNEX SPOT MONITOR

SERIAL No. : 100098421219

LOCATION : Heart Center

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

CALIBRATED DATE : 22 APRIL 2024 ISSUE DATE : 30 APRIL 2024

upatcha	

Performed by : Approved by : PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEVitalsign SimulatorFlukeProSim8254103123MD7777 Jun 2024

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PTP_0104240217

Page 2 of 3

Calibration Report

EQUIPMENT : MONITORS, PHYSIOLOGIC, VITAL SIGNS

ID CODE : PTP04523

MANUFACTURER : WELCH ALLYN

MODEL : CONNEX SPOT MONITOR

SERIAL No. : 100098421219

DATE OF CALIBRATION : 22 APRIL 2024

ENVIRONMENT : TEMPERATURE 24 °C

RELATIVE HUMIDITY 56 %

PROCEDURE USED:

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

Systolic					
Standard Setting	UUC* Reading	Error	% Error	Uncertainty	
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)	
60	63.00	3.00	5.00	± 0.9238	
120	121.00	1.00	0.83	± 1.5628	
200	204.50	4.50	2.25	± 2.0600	

Diastolic					
Standard Setting	UUC* Reading	Error	% Error	Uncertainty	
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)	
30	34.75	4.75	15.83	± 0.9425	
80	80.75	0.75	0.94	± 1.5544	
150	144.75	-5.25	-3.50	± 1.5276	

Heart Rate					
Standard Setting	UUC* Reading	Error	% Error	Uncertainty	
(BPM)	(BPM)	(BPM)	(%)	(BPM)	
60	60.00	0.00	0.00	± 0.0150	
80	80.00	0.00	0.00	± 0.0460	
120	120.00	0.00	0.00	± 0.0460	



Certificate No. 24PTP_0104240217

Page 3 of 3

Calibration Report

EQUIPMENT : MONITORS, PHYSIOLOGIC, VITAL SIGNS

ID CODE : PTP04523

SERIAL No. : 100098421219

%Spo2					
Standard Setting	UUC* Reading	Error	% Error	Uncertainty	
(%SPO ₂)	(%SPO ₂)	(%SPO ₂)	(%)	(%SPO ₂)	
90	90.00	0.00	0.00	± 0.1039	
96	96.00	0.00	0.00	± 0.1109	
100	100.00	0.00	0.00	± 0.1155	

UUC*: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-007/1 Rev.03 Page 3/3 Issued Date 20/07/2023