Certificate of Calibration

EQUIPMENT	:	OXIMETERS, PULSE
ID CODE	:	PYT3_01332
MANUFACTURER	:	MASIMO
MODEL	:	RADICAL-7
SERIAL No.	:	T040889
LOCATION	:	NURSERY
SUBMITTED BY	:	PHYATHAI 3 HOSPITAL
		111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160
		Tel : (662) 467-1111 Fax : (622) 467-1111
CALIBRATED DATE	:	16 DECEMBER 2024
ISSUE DATE	:	30 DECEMBER 2024

Performed by : _____   Approved by : _____

           SOM KONKAEW   *PUKARIN TONGKLIANG*

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTER	MANUFACTURER	MODEL	SERIAL NO.	CERTIFICATE NO.	DUE DATE
Tester, Vital Sign Simulator	FLUKE	PROSIM 8	5702059	24MD545	30 Apr 2025

2. THIS CERTIFICATION IS TRACEABLE TO :

   - Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY

Health

Calibration Report

EQUIPMENT   :   OXIMETERS, PULSE

ID CODE   :   PYT3_01332

MANUFACTURER   :   MASIMO

MODEL   :   RADICAL-7

SERIAL No.   :   T040889

DATE OF CALIBRATION   :   16 DECEMBER 2024

ENVIRONMENT   :   TEMPERATURE   25 °C

                            RELATIVE HUMIDITY   55 %

PROCEDURE USED:

   This instrument was calibration by comparison with standard

MEASURMENT RESULT:

[/] Without Adjustment   [ ] Before Adjustment   [ ] After Adjustment

%Spo2				
Standard Setting ($\%SPO_2$)	UUC* Reading ($\%SPO_2$)	Error ($\%SPO_2$)	% Error (%)	Uncertainty ($\%SPO_2$)
90	90.01	0.01	0.01	± 0.1106
96	96.02	0.02	0.02	± 0.0734
100	100.01	0.01	0.01	± 0.0487

Heart Rate				
Standard Setting (BPM)	UUC* Reading (BPM)	Error (BPM)	% Error (%)	Uncertainty (BPM)
60	60.00	-0.00	-0.00	± 0.0187
80	79.97	-0.03	-0.04	± 0.1160
120	120.00	-0.00	-0.00	± 0.1131

UUC* : Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%