

Report No.

24PTP_01PM04240291

Page 1 of 2

Preventive Maintenance Action Report

EQUIPMENT : MODULE, ECG/SPO2/NIBP/IBP/TEMP

ID CODE : PTP00616

MANUFACTURER : NIHON KOHDEN

MODEL : AY-663P

SERIAL No. : 6326

LOCATION : Operating Room

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

PM. DATE : 24 APRIL 2024

ISSUE DATE : 3 MAY 2024

Performed by:

Approved by:

5. 0.

USANEE CHALONGPHOKSILCHAI PHUWASIN YIWKIM

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE

2. THIS CERTIFICATION IS TRACEABLE TO:

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Report No. 24PTP_01PM04240291

Page 2 of 2

Preventive Maintenance Action Report

EQUIPMENT : MODULE, ECG/SPO2/NIBP/IBP/TEMP

ID CODE : PTP00616

SERIAL No. : 6326

1.Report

No.	List	Pass	Fail	None	Comment
1	This Module Calibrate With Monitor	/			

2.Remark MODULE OF ID Code :PTP00664

Electrical Safety Test: Master used Electrical Safety Analyzer (Metron / QA-90) according to IEC.60601.1

FI-BME-NHS-CP-001/3 Rev.09 Page 2/2 Issued Date 20/07/2023