

Certificate No.

24PYT3_DEC12240065

Page 1 of 3

Certificate of Calibration

EQUIPMENT : MONITORS, BEDSIDE, PHYSIOLOGIC

ID CODE : PYT3_00445

MANUFACTURER : PHILIPS

MODEL : INTELLIVUE MP30(M8002A)

SERIAL No. : DE72892456

LOCATION : LABOR ROOM(ห้องคลอด)

SUBMITTED BY : PHYATHAI 3 HOSPITAL

111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160

Tel: (662) 467-1111 Fax: (622) 467-1111

CALIBRATED DATE : 1 DECEMBER 2024

ISSUE DATE : 30 DECEMBER 2024

Performed by : Approved by : PUKARIN TONGKLIANG

SOM KONKAEW PUKARIN TONGKLIANG

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATETester, Vital SignFLUKEPROSIM 8570205924MD54530 Apr 2025

Simulator

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PYT3_DEC12240065

Page 2 of 3

Calibration Report

EQUIPMENT : MONITORS, BEDSIDE, PHYSIOLOGIC

ID CODE : PYT3_00445

MANUFACTURER : PHILIPS

MODEL : INTELLIVUE MP30(M8002A)

SERIAL No. : DE72892456

DATE OF CALIBRATION : 1 DECEMBER 2024

ENVIRONMENT : TEMPERATURE 25 °C

> RELATIVE HUMIDITY 55 %

PROCEDURE USED:

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

Systolic						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
60	59.98	-0.02	-0.04	± 0.0515		
120	121.01	1.01	0.84	± 0.0865		
200	194.96	-5.04	-2.52	± 0.0557		

Diastolic						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
30	30.99	0.99	3.31	± 0.0890		
80	81.01	1.01	1.26	± 0.0984		
150	143.02	-6.98	-4.66	± 0.0734		

Heart Rate						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(BPM)	(BPM)	(BPM)	(%)	(BPM)		
60	60.02	0.02	0.03	± 0.0363		
80	79.99	-0.01	-0.02	± 0.1116		
120	120.00	0.00	0.00	± 0.1100		

FI-BME-NHS-CP-007/1 Rev.04 Page 2/3 Issued Date 20/07/2024



Certificate No. 24PYT3_DEC12240065

Page 3 of 3

Calibration Report

EQUIPMENT : MONITORS, BEDSIDE, PHYSIOLOGIC

ID CODE : PYT3_00445

SERIAL No. : DE72892456

%Spo2						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(%SPO ₂)	(%SPO ₂)	(%SPO ₂)	(%)	(%SPO ₂)		
90	90.99	0.99	1.10	± 0.0752		
96	96.99	0.99	1.03	± 0.1076		
100	100.04	0.04	0.04	± 0.0639		

UUC*: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-007/1 Rev.04 Page 3/3 Issued Date 20/07/2024