

Certificate No.

24PTP_0104240236

Page 1 of 3

Certificate of Calibration

EQUIPMENT : MONITORS, PHYSIOLOGIC, VITAL SIGNS

ID CODE : PTP03789

MANUFACTURER : WELCH ALLYN

MODEL : CONNEX SPOT MONITOR

SERIAL No. : 100073772020

LOCATION : Patient Wards 408

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

CALIBRATED DATE : 23 APRIL 2024 ISSUE DATE : 30 APRIL 2024

	Usanee
Performed by:	

3,1024

USANEE CHALONGPHOKSILCHAI

PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

Approved by:

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE 7 Jun 2024 Vitalsign Simulator Fluke ProSim8 2541031 23MD777 ATC-140A 12 Jul 2024 Temperature Ametex 561460-00403 WB23062779 Calibrator (Source)

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)
- N-Health

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PTP_0104240236

Page 2 of 3

Calibration Report

EQUIPMENT : MONITORS, PHYSIOLOGIC, VITAL SIGNS

ID CODE : PTP03789

MANUFACTURER : WELCH ALLYN

MODEL : CONNEX SPOT MONITOR

SERIAL No. : 100073772020

DATE OF CALIBRATION : 23 APRIL 2024

ENVIRONMENT : TEMPERATURE 25 °C

RELATIVE HUMIDITY 55 %

PROCEDURE USED:

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

Systolic				
Standard Setting	UUC* Reading	Error	% Error	Uncertainty
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)
60	58.00	-2.00	-3.33	± 0.9238
120	114.00	-6.00	-5.00	± 1.2702
200	205.00	5.00	2.50	± 1.7321

Diastolic					
Standard Setting	UUC* Reading	Error	% Error	Uncertainty	
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)	
30	36.00	6.00	20.00	± 0.7506	
80	82.00	2.00	2.50	± 1.0393	
150	145.00	-5.00	-3.33	± 1.4434	

Heart Rate					
Standard Setting	UUC* Reading	Error	% Error	Uncertainty	
(BPM)	(BPM)	(BPM)	(%)	(BPM)	
60	60.00	0.00	0.00	± 0.0150	
80	80.00	0.00	0.00	± 0.0460	
120	120.00	0.00	0.00	± 0.0460	



Certificate No. 24PTP_0104240236

Page 3 of 3

Calibration Report

EQUIPMENT : MONITORS, PHYSIOLOGIC, VITAL SIGNS

ID CODE : PTP03789

SERIAL No. : 100073772020

%Spo2				
Standard Setting	UUC* Reading	Error	% Error	Uncertainty
(%SPO ₂)	(%SPO ₂)	(%SPO ₂)	(%)	(%SPO ₂)
90	90.00	0.00	0.00	± 0.1039
96	96.00	0.00	0.00	± 0.1109
100	100.00	0.00	0.00	± 0.1155

Temperature				
UUC* Setting	Standard Reading	Error	% Error	Uncertainty
(°C)	(°C)	(°C)	(%)	(°C)
35	35.10	0.10	0.29	± 0.3100
37	37.10	0.10	0.27	± 0.3100
39	39.00	0.00	0.00	± 0.3100

UUC*: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-007/1 Rev.03 Page 3/3 Issued Date 20/07/2023