

Certificate No.

24PTP_0104240272

Page 1 of 3

Certificate of Calibration

EQUIPMENT : MONITORS, PHYSIOLOGIC, VITAL SIGNS

ID CODE : PTP01350

MANUFACTURER : CRITICARE

MODEL : 506DN

SERIAL No. : 113054821

LOCATION : Patient Wards 15

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

CALIBRATED DATE : 24 APRIL 2024

ISSUE DATE : 1 MAY 2024

Kanittha.

KANITTHA SRIPRAPAD PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

Approved by:

CONDITION OF THIS RESULT OF TEST

Performed by :

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEVitalsign SimulatorFlukeProSim8254103123MD7777 Jun 2024

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PTP_0104240272

Page 2 of 3

Calibration Report

EQUIPMENT : MONITORS, PHYSIOLOGIC, VITAL SIGNS

ID CODE : PTP01350

MANUFACTURER : CRITICARE

MODEL : 506DN

SERIAL No. : 113054821

DATE OF CALIBRATION : 24 APRIL 2024

ENVIRONMENT : TEMPERATURE 25 °C

RELATIVE HUMIDITY 55 %

PROCEDURE USED:

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

Systolic						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
60	59.00	-1.00	-1.67	± 0.9238		
120	116.00	-4.00	-3.33	± 1.2702		
200	194.50	-5.50	-2.75	± 1.8258		

Diastolic						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(mmHg)	(mmHg)	(mmHg)	(%)	(mmHg)		
30	34.00	4.00	13.33	± 0.7506		
80	81.00	1.00	1.25	± 1.0393		
150	154.00	4.00	2.67	± 1.4434		

Heart Rate						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(BPM)	(BPM)	(BPM)	(%)	(BPM)		
60	60.00	0.00	0.00	± 0.0150		
80	80.00	0.00	0.00	± 0.0460		
120	120.00	0.00	0.00	± 0.0460		

FI-BME-NHS-CP-007/1 Rev.03 Page 2/3 Issued Date 20/07/2023



Certificate No. 24PTP_0104240272

Page 3 of 3

Calibration Report

EQUIPMENT : MONITORS, PHYSIOLOGIC, VITAL SIGNS

ID CODE : PTP01350 SERIAL No. : 113054821

%Spo2						
Standard Setting	UUC* Reading	Error	% Error	Uncertainty		
(%SPO ₂)	(%SPO ₂)	(%SPO ₂)	(%)	(%SPO ₂)		
90	90.00	0.00	0.00	± 0.1039		
96	96.00	0.00	0.00	± 0.1109		
100	100.00	0.00	0.00	± 0.1155		

UUC*: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-007/1 Rev.03 Page 3/3 Issued Date 20/07/2023