

Report No. 24PYT3\_DECPM12240052

Page 1 of 3

# **Preventive Maintenance Action Report**

EQUIPMENT : MONITORS, BEDSIDE, PHYSIOLOGIC

ID CODE : PYT3\_06721

MANUFACTURER : PHILIPS

MODEL : INTELLIVUE MP5(M8105A)

SERIAL No. : DE656B9982

LOCATION : NURSERY(เด็กอ่อน)

SUBMITTED BY : PHYATHAI 3 HOSPITAL

111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160

Tel: (662) 467-1111 Fax: (622) 467-1111

PM. DATE : 1 DECEMBER 2024
ISSUE DATE : 30 DECEMBER 2024

Performed by: Approved by:

SOM KONKAEW

PUKARIN TONGKLIANG

### **CONDITION OF THIS RESULT OF TEST**

#### 1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEELECTRICALFLUKEESA6142421039L202403400-000730 May 2025

SAFETY ANALYZER

#### 2. THIS CERTIFICATION IS TRACEABLE TO:

 $- Temperature, Dimension, mass, Anemometer, Electrical/Frequency, Relative\ Humidity and the property of the$ 

# 3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



*Report No.* 24PYT3\_DECPM12240052

Page 2 of 3

# **Preventive Maintenance Action Report**

EQUIPMENT : MONITORS, BEDSIDE, PHYSIOLOGIC

ID CODE : PYT3\_06721 SERIAL No. : DE656B9982

## 1. Qualitative Tests

No.	List	Pass	Fail	None	Comment
1	Chassis/Housing	/			
2	Mount/Fasteners	/			
3	Casters/Brakes	/			
4	AC Plug	/			
5	Line Cord	/			
6	Strain Reliefs	/			
7	Circuit Breaker/Fuse	/			
8	Tubes/Bulbs	/			
9	Fittings/Connectors	/			
10	Transducers	/			
11	Controls/Switches	/			
12	Battery/Charger	/			
13	Time/Date Settings	/			
14	Cybersecurity			/	
15	User Calibration/Self-Test			/	
16	Alarms	/			
17	Labeling	/			
18	Accessories	/			

### 2. Quantitative Tests

No.	List	Pass	Fail	None	Comment
1	Grounding Resistance ≤0.5 Ω	/			0.380 Ω
2	Chassis Leakage Current ≤500 µA	/			212 μΑ
3	Air Leakage ≤15 mmHg/min	/			
4	Static Pressure Accuracy ±3 mmHg or	/			
	*Oscillometric Pressure ±8 mmHg				
5	Maximum Pressure (NIBP) ≤330 mmHg	/	·		
6	Heart Rate ±10%	/			

### 3. Preventive Maintenance

No.	List	Pass	Fail	None	Comment
1	Clean the exterior and accessories	/			
2	Lubricate			/	



*Report No.* 24PYT3\_DECPM12240052

Page 3 of 3

# **Preventive Maintenance Action Report**

EQUIPMENT : MONITORS, BEDSIDE, PHYSIOLOGIC

ID CODE : PYT3\_06721 SERIAL No. : DE656B9982

### 3. Preventive Maintenance (Continues)

	No.	List	Pass	Fail	None	Comment
	3	Adjust			/	
ſ	4	Replace			/	

FI-BME-NHS-CP-007/3 Rev.04 Page 3/3 Issued Date 20/07/2024