

Report No.

24PYT3\_DECPM12240020

Page 1 of 3

# **Preventive Maintenance Action Report**

	SOM KONKAEW	PUKARIN TONGKLIANG
Performed by :	Gfm	Approved by :
ISSUE DATE	: 30 DECEMI	3ER 2024
PM. DATE	: 1 DECEMB	ER 2024
	Tel: (662) 4	67-1111 Fax : (622) 467-1111
	111 Phet Kasen	n Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 1016
SUBMITTED BY	: PHYATHAI	3 HOSPITAL
LOCATION	: WARD 15	
SERIAL No.	: CN2120751	1
MODEL	: SURESIGNS	S VSI
MANUFACTURER	: PHILIPS	
ID CODE	: PYT3_0203	6
EQUIPMENT	: MONITORS	S, PATIENT TRANSPORT

#### **CONDITION OF THIS RESULT OF TEST**

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE

2. THIS CERTIFICATION IS TRACEABLE TO:

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



*Report No.* 24PYT3\_DECPM12240020

Page 2 of 3

## **Preventive Maintenance Action Report**

EQUIPMENT : MONITORS, PATIENT TRANSPORT

ID CODE : PYT3\_02036 SERIAL No. : CN21207511

### 1. Qualitative Tests

No.	List	Pass	Fail	None	Comment
1	Chassis/Housing	/			
2	Mount/Fasteners	/			
3	Casters/Brakes	/			
4	AC Plug	/			
5	Line Cord	/			
6	Strain Reliefs	/			
7	Circuit Breaker/Fuse	/			
8	Tubes/Bulbs	/			
9	Fittings/Connectors	/			
10	Transducers	/			
11	Controls/Switches	/			
12	Battery/Charger	/			
13	Time/Date Settings	/			
14	Cybersecurity			/	
15	User Calibration/Self-Test			/	
16	Alarms	/			
17	Labeling	/			
18	Accessories	/			

#### 2. Quantitative Tests

No.	List	Pass	Fail	None	Comment
1	1 Grounding Resistance ≤0.5 Ω				0.30 Ω
2	Chassis Leakage Current ≤500 µA	/			28.1 μΑ
3	Air Leakage ≤15 mmHg/min	/			
4	Static Pressure Accuracy ±3 mmHg or	/			
	*Oscillometric Pressure ±8 mmHg				
5	Maximum Pressure (NIBP) ≤330 mmHg	/	·		
6	Heart Rate ±10%	/			

#### 3. Preventive Maintenance

No.	List	Pass	Fail	None	Comment
1	Clean the exterior and accessories	/			
2	Lubricate			/	



*Report No.* 24PYT3\_DECPM12240020

Page 3 of 3

# **Preventive Maintenance Action Report**

EQUIPMENT : MONITORS, PATIENT TRANSPORT

ID CODE : PYT3\_02036 SERIAL No. : CN21207511

### **3. Preventive Maintenance (Continues)**

No.	List	Pass	Fail	None	Comment
3	Adjust			/	
4	Replace			/	

FI-BME-NHS-CP-007/3 Rev.04 Page 3/3 Issued Date 20/07/2024