

Certificate No.

24PTP_0104240021

Page 1 of 2

Certificate of Calibration

EQUIPMENT : INFUSION PUMPS

ID CODE : PTP02653

MANUFACTURER : TOP

MODEL : TOP-2200

SERIAL No. : ELZ8607S

LOCATION : Patient Wards 406

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

CALIBRATED DATE : 16 APRIL 2024

ISSUE DATE : 3 MAY 2024

Supatcha

Performed by : _____ Approved by :

21004

SUPATCHA BUAPUEAN PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEInfusion PumpRIGEL MedicalMulti-Flo23Q-141223MD116924 Aug 2024

Tester

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PTP_0104240021

Page 2 of 2

CODE	: INI	FUSION PUMPS			
	: PT	P02653			
IANUFACTURER	: ТО	P			
IODEL	: ТО	P-2200			
ERIAL No.	: EL	Z8607S			
ATE OF CALIBRA		: 16 APRIL 2024			
NVIRONMENT		MPERATURE	24 °C		
NVIKONWIENI		LATIVE HUMIDITY	56 %		
/ Without Adjus	stment	Before Adjustment	After	Adjustmen	nt
UUC* Setting	Standard Reading	Error	% Error	Unc	ertainty
(ml/hr)	(ml/hr)	(ml/hr)	(%)		nl/hr)
5	4.78	-0.22	-4.35	± (0.0650
80	78.36	-1.64	-2.05	± (0.2634
200	193.25	-6.75	-3.38	± (0.7138
Occlusion Pressur	re < 1000 mmHg]			
UUC* Setting	Standard Reading				
(mmHg)	(mmHg)				
1000	366.00				
		Drops/	/ml.	20	ml.