

Certificate No.

24PYT3\_DEC12240029

Page 1 of 2

### **Certificate of Calibration**

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PYT3\_05490

MANUFACTURER : EDAN

SERIAL No. : -

**MODEL** 

LOCATION : WARD 11

SUBMITTED BY : PHYATHAI 3 HOSPITAL

H100B

111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160

Tel: (662) 467-1111 Fax: (622) 467-1111

CALIBRATED DATE : 1 DECEMBER 2024
ISSUE DATE : 30 DECEMBER 2024

In

Performed by:		Approved by :	1		
	SOM KONKAFW		PUKARIN TONGKI IANG		

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

### **CONDITION OF THIS RESULT OF TEST**

#### 1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATETester, Vital SignFLUKEPROSIM 8570205924MD54530 Apr 2025Simulator

### 2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

## 3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PYT3\_DEC12240029

Page 2 of 2

# **Calibration Report**

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PYT3\_05490

MANUFACTURER : EDAN MODEL : H100B

SERIAL No. : -

DATE OF CALIBRATION : 1 DECEMBER 2024

ENVIRONMENT : TEMPERATURE 25 °C

RELATIVE HUMIDITY 55 %

#### **PROCEDURE USED:**

This instrument was calibration by comparison with standard

### **MEASURMENT RESULT:**

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

%Spo2				
Standard Setting	UUC* Reading	Error	% Error	Uncertainty
(%SPO <sub>2</sub> )	(%SPO <sub>2</sub> )	(%SPO <sub>2</sub> )	(%)	(%SPO <sub>2</sub> )
90	91.05	1.05	1.16	± 0.0595
96	97.00	1.00	1.04	± 0.0946
100	99.98	-0.02	-0.02	± 0.1016

Heart Rate					
Standard Setting	UUC* Reading	Error	% Error	Uncertainty	
(BPM)	(BPM)	(BPM)	(%)	(BPM)	
60	60.02	0.02	0.03	± 0.0363	
80	80.01	0.01	0.01	± 0.1143	
120	120.01	0.00	0.00	± 0.1131	

### **UUC\*: Unit Under Calibration**

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-012/1 Rev.04 Page 2/2 Issued Date 20/07/2024