

Certificate No.

25PYT3_0101250288

Page 1 of 3

Certificate of Calibration

EQUIPMENT : ANESTHESIA UNITS

ID CODE : PYT3_04922

MANUFACTURER : GE HEALTHCARE

MODEL : AVANCE CS2

SERIAL No. : APKV01317

LOCATION : ANESTHESIA

SUBMITTED BY : PHYATHAI 3 HOSPITAL

111 Phet Kasem Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 10160

Tel: (662) 467-1111 Fax: (622) 467-1111

CALIBRATED DATE : 13 JANUARY 2025 ISSUE DATE : 25 JANUARY 2025

Performed by : Approved by : SUPATCHA BUAPUEAN PA

PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE 14 Jun 2025 Flowmeter Tester Drycal Defender 510 118211 24MD826 Gas Flow Analyzer Fluke Biomedical 24MD829 14 Jun 2025 VT650 6069725

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)
- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 25PYT3_0101250288

Page 2 of 3

Calibration Report

EQUIPMENT : ANESTHESIA UNITS

ID CODE : PYT3_04922

MANUFACTURER : GE HEALTHCARE

MODEL : AVANCE CS2

SERIAL No. : APKV01317

DATE OF CALIBRATION : 13 JANUARY 2025

ENVIRONMENT : TEMPERATURE 23 °C

RELATIVE HUMIDITY 50 %

PROCEDURE USED:

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

OXYGEN FLOW RATE				
UUC* Setting	Standard Reading	Error	% Error	Uncertainty
(LPM)	(LPM)	(LPM)	(%)	(LPM)
2	1.93	-0.07	-3.50	± 0.0512
5	4.88	-0.12	-2.35	± 0.0667
7	6.83	-0.17	-2.39	± 0.1026

N2O FLOW RATE (MIXED O2:N2O = 1:1)					
UUC* Setting	Standard Reading	Error	% Error	Uncertainty	
(LPM)	(LPM)	(LPM)	(%)	(LPM)	
4	3.85	-0.15	-3.69	± 0.0552	
10	9.82	-0.18	-1.83	± 0.1228	
14	13.86	-0.14	-1.00	± 0.1529	

AIR FLOW RATE					
UUC* Setting	Standard Reading	Error	% Error	Uncertainty	
(LPM)	(LPM)	(LPM)	(%)	(LPM)	
2	1.93	-0.07	-3.63	± 0.0339	
5	4.90	-0.10	-2.00	± 0.0694	
7	6.95	-0.06	-0.79	± 0.0959	

FI-BME-NHS-CP-022/1 Rev.04 Page 2/3 Issued Date 20/07/2024



Certificate No. 25PYT3_0101250288

Page 3 of 3

Calibration Report

EQUIPMENT : ANESTHESIA UNITS

ID CODE : PYT3_04922 SERIAL No. : APKV01317

PRESSURE					
UUC* Setting	Standard Reading	Error	% Error	Uncertainty	
(cmH ₂ O)	(cmH ₂ O)	(cmH ₂ O)	(%)	(cmH ₂ O)	
20	20.21	0.21	1.04	± 0.3366	
30	30.24	0.24	0.79	± 0.3377	
40	39.85	-0.15	-0.38	± 0.3367	

Note: LOCATION: OPERATING ROOM 4

UUC*: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

End of Report