

MANUFACTURER

Certificate No. 24PTP_0104240079

Page 1 of 2

Certificate of Calibration

EQUIPMENT : WARMING UNITS, PATIENT, RADIANT, INFANT

SUTINDA

ID CODE : PTP01072

MODEL : -

SERIAL No. : -

LOCATION : Nursery

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

CALIBRATED DATE : 17 APRIL 2024

ISSUE DATE : 3 MAY 2024

Performed by:

Approved by :

USANEE CHALONGPHOKSILCHAI

PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATETemperature TesterTESTO440 dP83962413DT2401035424 Jan 2025

2. THIS CERTIFICATION IS TRACEABLE TO:

- N-Health

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PTP_0104240079

Page 2 of 2

Calibration Report

EQUIPMENT : WARMING UNITS, PATIENT, RADIANT, INFANT

ID CODE : PTP01072

MANUFACTURER : SUTINDA

MODEL : SERIAL No. : -

DATE OF CALIBRATION : 17 APRIL 2024

ENVIRONMENT : TEMPERATURE 25 °C

RELATIVE HUMIDITY 55 %

PROCEDURE USED:

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

Hood Air Temperature								
UUC* Setting	Standard Reading	Error	% Error	Uncertainty				
(°C)	(°C)	(°C)	(%)	(°C)				
30	30.20	0.20	0.67	± 0.3055				
34	34.10	0.10	0.29	± 0.3055				
37	37.10	0.10	0.27	± 0.3055				

UUC*: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k = 2, providing a level of confidence of approximately 95%

2301/2 New Petchburi Soi 47 (soonvijai), Bangkapi, Huaykwang, Bangkok 10310 Tel: (66) 2762 4000 Fax: (66) 2762 4072

FI-BME-NHS-CP-010/1 Rev.03 Page 2/2 Issued Date 20/07/2023