

Report No.

24PYT3\_DECPM12240111

Page 1 of 2

# **Preventive Maintenance Action Report**

	SOM KONKAEW	PUKARIN TONGKLIANG
Performed by :	Gfm	Approved by :
ISSUE DATE	: 30 DECEMI	BER 2024
PM. DATE	: 3 DECEMB	ER 2024
	Tel: (662) 4	67-1111 Fax : (622) 467-1111
	111 Phet Kasen	n Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 1016
SUBMITTED BY	: PHYATHAI	3 HOSPITAL
LOCATION	: EENT( หูตาค	าอามูก)
SERIAL No.	: PC66-019	
MODEL	: M730	
MANUFACTURER	: SOLIDA	
ID CODE	: PYT3_0893	9
EQUIPMENT	: ASPIRATO	RS, EMERGENCY

#### **CONDITION OF THIS RESULT OF TEST**

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE

2. THIS CERTIFICATION IS TRACEABLE TO:

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Report No. 24PYT3\_DECPM12240111

Page 2 of 2

# **Preventive Maintenance Action Report**

EQUIPMENT : ASPIRATORS, EMERGENCY

ID CODE : PYT3\_08939

SERIAL No. : PC66-019

### 1. Qualitative Tests

No.	List	Pass	Fail	None	Comment
1	1 Chassis/Housing				
2	2 Tubes/Hoses				
3 Fittings/Connectors		/			
4	Filters			/	
5	5 Controls/Switches				
6	6 Overflow Protection			/	
7	7 Intermittent Operation			/	
8	8 Indicators/Display				
9	9 Labeling				
10	0 Accessories				

## 2. Quantitative Tests

No.	List	Pass	Fail	None	Comment
1	Maximum Free Flow surgical/tracheal ≥25 L/min,			/	
	thoracic ≥20 L/min				
2	Maximum Vacuum Low-volume>100 mmHg;			/	
	thoracic >50 mmHg; surgical/tracheal ≥400 mmHg				
3	Vacuum Gauge Accuracy ±5% FSO(Full Scale	/			
	output)				

#### 3. Preventive Maintenance

No.	List	Pass	Fail	None	Comment
1	Clean the exterior and accessories	/			
2	Lubricate			/	
3	Adjust			/	
4	Replace			/	