

Report No.

24PYT3_DECPM12240034

Page 1 of 2

Preventive Maintenance Action Report

| EQUIPMENT | : OXIMETER | RS, PULSE | | | | |
|----------------|----------------|---|--|--|--|--|
| ID CODE | : PYT3_0929 | PYT3_09295 | | | | |
| MANUFACTURER | : BEIJING C | HOICE | | | | |
| MODEL | : MD300C1 | | | | | |
| SERIAL No. | : - | | | | | |
| LOCATION | : WARD 8 | | | | | |
| SUBMITTED BY | : PHYATHA | I 3 HOSPITAL | | | | |
| | 111 Phet Kaser | n Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 1016 | | | | |
| | Tel: (662) 4 | 467-1111 Fax : (622) 467-1111 | | | | |
| PM. DATE | : 12 DECEM | BER 2024 | | | | |
| ISSUE DATE | : 30 DECEM | BER 2024 | | | | |
| | 2: | | | | | |
| Darformed by | Gen | Approved by | | | | |
| Performed by : | SOM KONKAEW | Approved by : PUKARIN TONGKLIANG | | | | |

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE

2. THIS CERTIFICATION IS TRACEABLE TO:

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Report No. 24PYT3_DECPM12240034

Page 2 of 2

Preventive Maintenance Action Report

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PYT3_09295

SERIAL No. : -

1. Qualitative Tests

| No. | List | Pass | Fail | None | Comment |
|-----|----------------------------|------|------|------|---------|
| 1 | Chassis/Housing | / | | | |
| 2 | 2 Mount/Fasteners | | | / | |
| 3 | 3 AC Plug/Receptacles | | | / | |
| 4 | Line Cord | | | / | |
| 5 | Strain Reliefs | | | / | |
| 6 | Circuit Breaker/Fuse | | | / | |
| 7 | Cables | | | / | |
| 8 | Connectors | / | | | |
| 9 | Probes/Sensors/Transducers | / | | | |
| 10 | Controls/Switches | / | | | |
| 11 | Battery/Charger | / | | | |
| 12 | Indicators/Displays | / | | | |
| 13 | Self-Test | / | | | |
| 14 | Time/Date Settings | / | | | |
| 15 | Cybersecurity | | | / | |
| 16 | Alarms | / | | | |
| 17 | Audible Signals | / | | | |
| 18 | Labeling | / | | | |

2. Quantitative Tests

| No. | List | Pass | Fail | None | Comment |
|-----|---------------------------------|------|------|------|---------|
| 1 | Grounding Resistance ≤0.5 Ω | | | / | |
| 2 | Chassis Leakage Current ≤500 µA | | | / | |

3. Preventive Maintenance

| No. | List | Pass | Fail | None | Comment |
|-----|------------------------------------|------|------|------|---------|
| 1 | Clean the exterior and accessories | / | | | |
| 2 | Lubricate | | | / | |
| 3 | Adjust | | | / | |
| 4 | Replace | | | / | |