

Certificate No.

24PTP_0104240137

Page 1 of 2

Certificate of Calibration

EQUIPMENT : INFUSION PUMPS, ENTERAL FEEDING

ID CODE : PTP04529

MANUFACTURER : HAWKMED

MODEL : HK-300

SERIAL No. : 01163068

LOCATION : I.C.U

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

CALIBRATED DATE : 18 APRIL 2024

ISSUE DATE : 3 MAY 2024

Chakkaphan

Performed by : CHAKKAPHAN KHAMPHUMI

Approved by :

PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEInfusion PumpRIGEL MedicalMulti-Flo23Q-141223MD116924 Aug 2024

Tester

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PTP_0104240137

Page 2 of 2

Calibration Report

EQUIPMENT	: INI	FUSION PUMPS, EN	NTERAL FEEDING	
ID CODE	: PT	P04529		
MANUFACTURER	: НА	WKMED		
MODEL	: НК	C-300		
SERIAL No.	: 011	163068		
DATE OF CALIBRA		APRIL 2024		
ENVIRONMENT		MPERATURE	24 °C	
		LATIVE HUMIDITY	Y 55 %	
/ Without Adjus	SULT:	Before Adjustment	After A	Adjustment
		Before Adjustment	After A	Adjustment
Flow Rate UUC* Setting	stment	Before Adjustment Error	After A	
Flow Rate		,		
Flow Rate UUC* Setting	Standard Reading	Error	% Error	Uncertaint (ml/hr)
Flow Rate UUC* Setting (ml/hr)	Standard Reading (ml/hr)	Error (ml/hr)	% Error (%)	Uncertaint (ml/hr) ± 0.2600
Flow Rate UUC* Setting (ml/hr) 100	Standard Reading (ml/hr) 104.80	Error (ml/hr) 4.80	% Error (%) 4.80	Uncertaint
Flow Rate UUC* Setting (ml/hr) 100 200	Standard Reading (ml/hr) 104.80 201.78	Error (ml/hr) 4.80 1.78	% Error (%) 4.80 0.89	Uncertaint (ml/hr) ± 0.2600 ± 0.7100

2301/2 New Petchburi Soi 47 (soonvijai), Bangkapi, Huaykwang, Bangkok 10310 Tel: (66) 2762 4000 Fax: (66) 2762 4072

FI-BME-NHS-CP-072/1 Rev.03 Page 2/2 Issued Date 20/07/2023