

Report No.

24PTP_01PM04240072

Page 1 of 2

Preventive Maintenance Action Report

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PTP03803

MANUFACTURER : JUMPER

MODEL : JPD-500E

SERIAL No. : 8823500E408006

LOCATION : Patient Wards 406

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

PM. DATE : 17 APRIL 2024

USANEE CHALONGPHOKSILCHAI

ISSUE DATE : 1 MAY 2024

Performed by:

Approved by:

PHUWASIN YIWKIM

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE

2. THIS CERTIFICATION IS TRACEABLE TO:

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Report No. 24PTP_01PM04240072

Page 2 of 2

Preventive Maintenance Action Report

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PTP03803

SERIAL No. : 8823500E408006

1. Qualitative Tests

No.	List	Pass	Fail	None	Comment
1	Chassis/Housing	/			
2	Mount/Fasteners			/	
3	AC Plug/Receptacles			/	
4	Line Cord			/	
5	Strain Reliefs			/	
6	Circuit Breaker/Fuse			/	
7	Cables			/	
8	Connectors			/	
9	Probes/Sensors/Transducers	/			
10	Controls/Switches	/			
11	Battery/Charger	/			
12	Indicators/Displays	/			
13	Self-Test	/			
14	Time/Date Settings			/	
15	Cybersecurity			/	
16	Alarms	/			
17	Audible Signals	/			
18	Labeling	/			

2. Quantitative Tests

No.	List	Pass	Fail	None	Comment
1	Grounding Resistance ≤0.5 Ω			/	
2	Chassis Leakage Current ≤500 µA			/	

3. Preventive Maintenance

No.	List	Pass	Fail	None	Comment
1	Clean the exterior and accessories	/			
2	Lubricate			/	
3	Adjust			/	
4	Replace			/	

Electrical Safety Test: Master used Electrical Safety Analyzer (Metron / QA-90) according to IEC.60601.1