

Report No.

24PYT3_DECPM12240158

Page 1 of 2

Preventive Maintenance Action Report

	SOM KONKAEW	PUKARIN TONGKLIANG
Performed by :	Gfm	Approved by :
ISSUE DATE	: 30 DECEM	IBER 2024
PM. DATE	: 16 DECEM	IBER 2024
	Tel: (662) 4	467-1111 Fax: (622) 467-1111
	111 Phet Kaser	m Rd., Pak Khlong Phasi Charoen, Phasi Charoen Bangkok 1016
SUBMITTED BY	: РНҮАТНА	I 3 HOSPITAL
LOCATION	: NURSERY	
SERIAL No.	: T019604	
MODEL	: RADICAL-	7
MANUFACTURER	: MASIMO	
ID CODE	: PYT3_0174	44
EQUIPMENT	: OXIMETER	RS, PULSE

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE

2. THIS CERTIFICATION IS TRACEABLE TO:

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Report No. 24PYT3_DECPM12240158

Page 2 of 2

Preventive Maintenance Action Report

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PYT3_01744

SERIAL No. : T019604

1. Qualitative Tests

No.	List	Pass	Fail	None	Comment
1	Chassis/Housing	/			
2	Mount/Fasteners			/	
3	AC Plug/Receptacles			/	
4	Line Cord			/	
5	Strain Reliefs			/	
6	Circuit Breaker/Fuse			/	
7	Cables			/	
8	Connectors	/			
9	Probes/Sensors/Transducers	/			
10	Controls/Switches	/			
11	Battery/Charger	/			
12	Indicators/Displays	/			
13	Self-Test	/			
14	Time/Date Settings	/			
15	Cybersecurity			/	
16	Alarms	/			
17	Audible Signals	/			
18	Labeling	/			

2. Quantitative Tests

No.	List	Pass	Fail	None	Comment
1	Grounding Resistance ≤0.5 Ω	/			$0.420~\Omega$
2	Chassis Leakage Current ≤500 μA	/			206 μΑ

3. Preventive Maintenance

No.	List	Pass	Fail	None	Comment
1	Clean the exterior and accessories	/			
2	Lubricate			/	
3	Adjust			/	
4	Replace			/	