

Report No.

24PYT3\_DECPM12240151

Page 1 of 2

# **Preventive Maintenance Action Report**

	SOM KONKAFW		PUKARIN TONGKI IANG
Performed by :	Gfm	Approved by:	D-7-
ISSUE DATE	: 30 DE	CEMBER 2024	
PM. DATE	: 13 DEG	CEMBER 2024	
	Tel : (6	662) 467-1111 Fax : (6	522) 467-1111
	111 Phet	Kasem Rd., Pak Khlong Pha	asi Charoen, Phasi Charoen Bangkok 10160
SUBMITTED BY	: PHYA	THAI 3 HOSPITAL	
LOCATION	: OPD C	CARDIO(ศูนย์หัวใจ)	
SERIAL No.	: N4497	9	
MODEL	: RAD-5	5V	
MANUFACTURER	: MASIN	MO	
ID CODE	: PYT3_	_01768	
EQUIPMENT	: OXIM	ETERS, PULSE	

#### **CONDITION OF THIS RESULT OF TEST**

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE

2. THIS CERTIFICATION IS TRACEABLE TO:

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



*Report No.* 24PYT3\_DECPM12240151

Page 2 of 2

## **Preventive Maintenance Action Report**

EQUIPMENT : OXIMETERS, PULSE

ID CODE : PYT3\_01768

SERIAL No. : N44979

### 1. Qualitative Tests

No.	List	Pass	Fail	None	Comment
1	Chassis/Housing	/			
2	2 Mount/Fasteners			/	
3	3 AC Plug/Receptacles			/	
4	Line Cord			/	
5	Strain Reliefs			/	
6	Circuit Breaker/Fuse			/	
7	Cables			/	
8	Connectors	/			
9	Probes/Sensors/Transducers	/			
10	Controls/Switches	/			
11	Battery/Charger	/			
12	Indicators/Displays	/			
13	Self-Test	/			
14	Time/Date Settings	/			
15	Cybersecurity			/	
16	Alarms	/			
17	Audible Signals	/			
18	Labeling	/			

#### 2. Quantitative Tests

No.	List	Pass	Fail	None	Comment
1	Grounding Resistance ≤0.5 Ω			/	
2	Chassis Leakage Current ≤500 µA			/	

### 3. Preventive Maintenance

No.	List	Pass	Fail	None	Comment
1	Clean the exterior and accessories	/			
2	Lubricate			/	
3	Adjust			/	
4	Replace			/	