

Report No.

24PTP\_01PM04240158

Page 1 of 2

## **Preventive Maintenance Action Report**

EQUIPMENT : MODULE, ECG/SPO2/NIBP

ID CODE : PTP00472

MANUFACTURER : PHILIPS

MODEL : M3001A

SERIAL No. : DE5630FGR8

LOCATION : I.C.U

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

PM. DATE : 18 APRIL 2024

ISSUE DATE : 2 MAY 2024

Supatcha

Performed by : \_\_\_\_\_ Approved by : \_\_\_\_\_

SUPATCHA BUAPUEAN PHUWASIN YIWKIM

## **CONDITION OF THIS RESULT OF TEST**

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE

2. THIS CERTIFICATION IS TRACEABLE TO:

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Report No. 24PTP\_01PM04240158

Page 2 of 2

## **Preventive Maintenance Action Report**

EQUIPMENT : MODULE, ECG/SPO2/NIBP

ID CODE : PTP00472

SERIAL No. : DE5630FGR8

## 1.Report

No.	List	Pass	Fail	None	Comment
1	This Module Calibrate With Monitor	/			

2.Remark MODULE OF ID Code: PTP00380

 $Electrical\ Safety\ Test: Master\ used\ Electrical\ Safety\ Analyzer\ (\ Metron\ /\ QA-90\ )\ according\ to\ IEC.60601.1$ 

FI-BME-NHS-CP-001/3 Rev.09 Page 2/2 Issued Date 20/07/2023