

Report No. 24PTP_01PM04240317

Page 1 of 3

Preventive Maintenance Action Report

EQUIPMENT : WARMING/COOLING UNITS, PATIENT, CIRCULATING-

LIQUID

ID CODE : PTP02531 MANUFACTURER : GAYMAR

MODEL : MEDI -THERM III (MTA7912)

SERIAL No. : G21004 LOCATION : I.C.U

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

PM. DATE : 25 APRIL 2024

ISSUE DATE : 1 MAY 2024

Performed by:

KANITTHA SRIPRAPAD

Approved by:

PHUWASIN YIWKIM

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTERMANUFACTURERMODELSERIAL NO.CERTIFICATE NO.DUE DATEElectrical SafetyFlukeESA615349403423MD7747 Jun 2024Analyzer

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Report No. 24PTP_01PM04240317

Page 2 of 3

Preventive Maintenance Action Report

EQUIPMENT : WARMING/COOLING UNITS, PATIENT, CIRCULATING-

LIQUID

ID CODE : PTP02531

SERIAL No. : G21004

1. Qualitative Tests

No.	List	Pass	Fail	None	Comment
1	Chassis/Housing	/			
2	Mount	/			
3	Casters/Brakes	/			
4	AC Plug	/			
5	Line Cord	/			
6	Strain Reliefs	/			
7	Circuit Breaker/Fuse	/			
8	Tubes/Hoses	/			
9	Fittings/Connectors	/			
10	Probes	/			
11	Filters	/			
12	Controls/Switches	/			
13	Heater	/			
14	Motor/Pump/Fan/Compressor	/			
15	Fluid Levels	/			
16	Indicators/Displays	/			
17	Self-Test	/			
18	Time/Date Settings			/	
19	Alarms/Interlocks	/			
20	Audible Signals	/			
21	Labeling	/			
22	Blankets	/			
23	Cybersecurity			/	

2. Quantitative Tests

_	_, <i>C</i>					
No.	List	Pass	Fail	None	Comment	
1	Grounding Resistance $\leq 0.5 \Omega$	/			0.334Ω	
2	Chassis Leakage Current ≤500 µA	/			130.7 μΑ	
3	High-Temperature Protection ≤43°C or	/				
	manufacturer specification					
4	Low Temperature Protection ±1°C of mfr spec	/				
5	Fluid Temperature Display ±1°C	/				
6	Patient Temperature Display and Probe ±1°C			/		



Report No. 24PTP_01PM04240317

Page 3 of 3

Preventive Maintenance Action Report

EQUIPMENT : WARMING/COOLING UNITS, PATIENT, CIRCULATING-

LIQUID

ID CODE : PTP02531

SERIAL No. : G21004

2. Quantitative Tests (Continues)

No.	List	Pass	Fail	None	Comment
7	Automatic Controller Switching ±0.5°C or mfr	/			
	spec				
8	Fluid Flow Unit alarms; mfr flow spec, if indicated	/			

3. Preventive Maintenance

No.	List	Pass	Fail	None	Comment
1	Clean the exterior and accessories	/			
2	Lubricate			/	
3	Adjust			/	
4	Replace			/	

Electrical Safety Test: Master used Electrical Safety Analyzer (Metron / QA-90) according to IEC.60601.1

2301/2 New Petchburi Soi 47 (soonvijai), Bangkapi, Huaykwang, Bangkok 10310 Tel: (66) 2762 4000 Fax: (66) 2762 4072

FI-BME-NHS-CP-066/3 Rev.03 Page 3/3 Issued Date 20/07/2023