

Certificate No.

24PTP_0104240307

Page 1 of 2

Certificate of Calibration

EQUIPMENT : STERILIZING UNITS, STEAM, TABLETOP

ID CODE : PTP02123

MANUFACTURER : EURONDA

MODEL : E9 MED

SERIAL No. : EHD160782

LOCATION : Dental & Implant Center

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax: -

CALIBRATED DATE : 25 APRIL 2024

ISSUE DATE : 3 MAY 2024

upatcha		むりゃん
	Approved by :	

SUPATCHA BUAPUEAN PHUWASIN YIWKIM

This certificate may not be reproduced except in full unless permission for reproduction has been obtained in writing from the calibration.

CONDITION OF THIS RESULT OF TEST

Performed by:

1. REFERENCE STANDARD INSTRUMENT:

MASTER	MANUFACTURER	MODEL	SERIAL NO.	CERTIFICATE NO.	DUE DATE
Temperature Data	Omega	OM-CP-	R25518	23I672	19 Jun 2024
Logger		HITEMP140			
Pressure Data	Omega	OM-CP-	P95847	23P1417	9 May 2024
logger		PR140-NPT			

2. THIS CERTIFICATION IS TRACEABLE TO:

- Technology Promotion Association (Thai-Japan)
- Technology Promotion Association (Thai-Japan)

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Certificate No. 24PTP_0104240307

Page 2 of 2

Calibration Report

EQUIPMENT : STERILIZING UNITS, STEAM, TABLETOP

ID CODE : PTP02123

MANUFACTURER : EURONDA

MODEL : E9 MED

SERIAL No. : EHD160782

DATE OF CALIBRATION : 25 APRIL 2024

ENVIRONMENT : TEMPERATURE 24 °C

RELATIVE HUMIDITY 56 %

PROCEDURE USED:

This instrument was calibration by comparison with standard

MEASURMENT RESULT:

/	Without Adjustment		Before Adjustment		After Adjustment
---	--------------------	--	-------------------	--	------------------

Temperature						
UUC* Setting	Standard Reading	Error	% Error	Uncertainty		
(°C)	(°C)	(°C)	(%)	(°C)		
134	135.21	1.21	0.90	± 0.0660		

Pressure						
UUC* Setting	Standard Reading	Error	% Error	Uncertainty		
(bar)	(bar)	(bar)	(%)	(bar)		
2.07	2.09	0.02	1.05	± 0.0016		

UUC*: Unit Under Calibration

The reported uncertainty of measurement was based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%

FI-BME-NHS-CP-074/1 Rev.03 Page 2/2 Issued Date 20/07/2023