

Report No.

24PTP_01PM04240357

Page 1 of 2

Preventive Maintenance Action Report

EQUIPMENT : PACEMAKERS, CARDIAC, EXTERNAL

ID CODE : PTP02538

MANUFACTURER : ST. JUDE MEDICAL

MODEL : 3077

SERIAL No. : 1322206

LOCATION : I.C.U

SUBMITTED BY : PHYATHAI PHAHOLYOTHIN HOSPITAL

670/1 Phaholyothin Road, Samsen Nai, Phaya Thai, Bangkok 10400

Tel: 02-271-7000 Fax:-

PM. DATE : 26 APRIL 2024

KANITTHA SRIPRAPAD

ISSUE DATE : 1 MAY 2024

Reformed by:

Approved by :

PHUWASIN YIWKIM

CONDITION OF THIS RESULT OF TEST

1. REFERENCE STANDARD INSTRUMENT:

MASTER MANUFACTURER MODEL SERIAL NO. CERTIFICATE NO. DUE DATE

2. THIS CERTIFICATION IS TRACEABLE TO:

3. THIS RESULT OF TEST WAS FOUND ACCURATE AS SHOW ON DATE AND PLACE OF TEST ONLY



Report No. 24PTP_01PM04240357

Page 2 of 2

Preventive Maintenance Action Report

EQUIPMENT : PACEMAKERS, CARDIAC, EXTERNAL

ID CODE : PTP02538

SERIAL No. : 1322206

1. Qualitative Tests

No.	List	Pass	Fail	None	Comment
1	Chassis/Housing	/			
2	Cables	/			
3	Terminals	/			
4	Controls/Switches	/			
5	Indicators/Displays	/			
6	Time/Date Settings			/	
7	Cybersecurity			/	
8	Labeling	/			
9	Accessories	/			

2. Quantitative Tests

No.	List	Pass	Fail	None	Comment
1	Pulse Width 0.1 to 2.0 msec, typical; ±10% mfr			/	
	spec, user adjustable				
2	Atrial-Ventricular Delay ±10% mfr spec			/	
3	Rate Accuracy ±5%			/	
4	Amplitude Accuracy ±10%	/			
5	Demand-Mode Sensitivity 0.5-20 mV or mfr spec			/	

3. Preventive Maintenance

No.	List	Pass	Fail	None	Comment
1	Clean the exterior and accessories	/			
2	Lubricate			/	
3	Adjust			/	
4	Replace			/	

 $Electrical\ Safety\ Test: Master\ used\ Electrical\ Safety\ Analyzer\ (\ Metron\ /\ QA-90\)\ according\ to\ IEC.60601.1$