

# Performance of a Low-parasitic Frequency Domain Multiplexing Readout

Amy E. Lowitz<sup>1</sup>, Amy N. Bender<sup>2</sup>, P. Barry<sup>2</sup>, T. Cecil<sup>2</sup>, C. Chang<sup>1,2</sup>, R. Divan<sup>2</sup>, M. Dobbs<sup>3</sup>, A. Gilbert<sup>3</sup>, S. Kuhlmann<sup>2</sup>, M. Lisovenko<sup>2</sup>, J. Montgomery<sup>3</sup>, V. Novosad<sup>2</sup>, S. Padin<sup>1,2</sup>, J. Pearson<sup>2</sup>, G. Wang<sup>2</sup>, V. Yefremenko<sup>2</sup>, J. Zhang<sup>2</sup>


<sup>1</sup>University of Chicago, <sup>2</sup>Argonne National Laboratory, <sup>3</sup>McGill University

## Introduction

Frequency-domain multiplexing is a readout technique for transition-edge sensor bolometer arrays used in modern CMB experiments. Here we present design details and performance measurements for a **low-parasitic frequency-domain multiplexing readout**. In this ‘**modified frequency domain multiplexing**’ (MfMux) readout, low parasitic impedance reduces crosstalk between readout channels. This allows **higher multiplexing factors**, and enables operation of lower-resistance bolometers, which **decreases the contribution of readout noise to the total NEP** by decreasing the required voltage bias. The MfMux system has demonstrated warm overbiased noise comparable to SPT-3G readout hardware, as well as a factor of two improvement in parasitic resistance compared to SPT-3G hardware. Ongoing work seeks to further reduce parasitic impedances, leading to a lab demonstration of a readout with low-resistance bolometers.

## Design and Implementation


- Moving the SQUID from the 4 K stage to the 250 mK stage
- This reduces the wiring length between the SQUID and LC chip
- Reduced parasitic impedance
  - Reduces crosstalk
  - Enables operation of low-resistance bolometers
- Improves scalability
- Retained as much of the existing DfMux design as possible.
- This reaps the benefits of **reduced crosstalk, reduced parasitic resistance, and improved scalability, while retaining as much technological maturity as possible.**
- **Magnetic shielding** for the SQUID is provided by an aluminum box covered with six layers of Metglas. Additional pinning of any remaining stray magnetic fields is provided by a small Nb foil onto which the SQUID is glued.
- The prototype boards are sized so they can be a **drop-in replacement** for SPT-3G-style LC boards in any of the SPT-3G testbeds or, in principle, the telescope itself.



## Key benefits


Reducing the stray impedance between the LC chip and the SQUID offers several benefits:

- Decreased stray resistance enables **operation of low- $R_{\text{normal}}$  bolometers**.
  - This **reduces  $\text{NEP}_{\text{readout}}$**  because the bolometers can be operated with lower  $V_{\text{bias}}$ .
- Reduced crosstalk enables denser packing of bolometers in frequency space
  - This **allows for higher multiplexing factors and improved scalability.**
- Leverages much of the existing dfMux architecture, which has proven on-sky performance
- Drop-in replacement for 3G-style LC boards



## Next Steps

SQUID improvements:



NIST SA-13 SQUIDs work well for testing small numbers of prototype mfMux boards, however because of their **thermal dissipation**, putting hundreds of them (as required for ultra-large focal planes) on a cold stage cooled with a typical sorption refrigerator would not work. Either a lower thermal dissipation SQUID or a higher cooling power refrigerator (such as a dilution refrigerator) would be required. We hope to test new, low-thermal-dissipation SQUIDs in the future. **Shown above** is the estimated thermal dissipation from a typical SA-13 SQUID, calculated as the output voltage  $\times$  the current bias.

## Noise and Crosstalk Characterization

While noise and parasitic performance has been promising so far, more detailed study is required to further mature this technology. This will be a key focus of work in the immediate future


## Low-resistance bolometer integration

One of the key goals of this technology is to enable operation of **low-resistance bolometers** with a frequency multiplexed readout. We have already shown a significant reduction in parasitic resistance, which is a major step towards that goal. Our next step will be to test the mfMux hardware with a ‘detector wafer’ where the detectors have been replaced with resistors at a range of resistances. This test wafer has already been fabricated, and will be used for mfMux testing in the immediate future.


## Performance

(Below Left) Shown is a network analysis for one MfMux test comb.

(Below Right) Performance of the NIST SA-13 SQUIDs is similar at 3.8 K, 2.4 K and 257 mK, with small improvements in peak-to-peak voltage and transimpedance at lower temperatures. Shown is a v-phi curve for a representative SQUID at three temperatures.




(Below) We have **operated 57 (of a possible 58) bolometers** on a single comb simultaneously in the TES superconducting transition to a fractional resistance of  $0.2 \times R_{\text{normal}}$ . Shown are RP curves for three representative bolometers. The ‘bendback’ feature results from the parasitic resistance, which is significant compared to the bolometer resistance when very deep in the superconducting transition.



(Below Left) The warm overbiased **noise performance** of the MfMux system is comparable to the performance of the standard SPT-3G DfMux hardware in the same laboratory testbed (shown) and on-sky.

(Below Right) The median **parasitic resistance** of the MfMux prototype is half that of 3G DfMux hardware under the same laboratory testing conditions. Vertical lines indicate the medians.



## Selected References

Lowitz et al, “Digital frequency multiplexing with sub-Kelvin SQUIDs,” Proc. SPIE (2018)

Bender et al, “Digital frequency domain multiplexing readout electronics for the next generation of millimeter telescopes,” Proc. SPIE (2012)

Dobbs et al, “Frequency multiplexed superconducting quantum interference device readout of large bolometer arrays for cosmic microwave background measurements,” Rev. Sci. Instrum. (2012)

## Acknowledgments

Work at the University of Chicago is supported by the National Science Foundation through grant PLR-1248097. Work at Argonne National Lab is supported by UChicago Argonne LLC, Operator of Argonne National Laboratory (Argonne). Argonne, a U.S. Department of Energy Office of Science Laboratory, is operated under contract no. DE-AC02-06CH11357. The McGill authors acknowledge funding from the Natural Sciences and Engineering Research Council of Canada, Canadian Institute for Advanced Research, and the Fonds de recherche du Québec Nature et technologies.