Clinical compliance

model	protocol	SBP MAE	SBP SD	AAMI	BHS	DBP MAE	DBP SD	AAMI	BHS
Random Forest	leaky	7.47	10.46	FAIL	C	4.34	6.32	PASS	A
Random Forest	calbased	7.14	10.12	FAIL	C	4.11	6.16	PASS	A
Random Forest	calfree	12.76	16.26	FAIL	D	8.38	10.62	FAIL	D
Random Forest	aami	19.67	23.01	FAIL	D	12.24	13.97	FAIL	D
Gradient Boosting	leaky	10.29	13.20	FAIL	D	6.13	7.98	PASS	B
Gradient Boosting	calbased	10.20	13.08	FAIL	D	6.03	7.92	PASS	B
Gradient Boosting	calfree	14.63	18.40	FAIL	D	9.00	11.29	FAIL	D
Gradient Boosting	aami	17.44	22.28	FAIL	D	11.14	13.59	FAIL	D
1D-CNN	leaky	11.10	14.32	FAIL	D	7.04	9.22	FAIL	C
1D-CNN	calbased	9.89	13.05	FAIL	D	6.29	8.41	FAIL	B
1D-CNN	calfree	14.89	18.52	FAIL	D	9.04	11.29	FAIL	D
1D-CNN	aami	16.53	21.54	FAIL	D	11.20	13.78	FAIL	D
ResNet1D	leaky	6.07	8.88	FAIL	B	3.87	5.96	PASS	A
ResNet1D	calbased	6.17	8.94	FAIL	B	3.92	6.02	PASS	A
ResNet1D	calfree	13.94	17.85	FAIL	D	8.94	11.40	FAIL	D
ResNet1D	aami	17.70	21.79	FAIL	D	11.40	13.33	FAIL	D
Transformer	leaky	14.46	18.65	FAIL	D	8.15	10.52	FAIL	D
Transformer	calbased	13.80	17.97	FAIL	D	8.06	10.41	FAIL	D
Transformer	calfree	15.44	19.60	FAIL	D	8.47	10.78	FAIL	D
Transformer	aami	17.03	21.70	FAIL	D	10.15	13.31	FAIL	D

AAMI passes when the mean error is within ±5 mmHg and its SD within 8. BHS grades on the share of errors under 5 / 10 / 15 mmHg (A is best, D is a fail).