

Bland-Altman — ResNet1D. A downward tilt means the model is pulling every patient towards the population average.
leaky
calbased
calfree
aami
SBP predicted − true (mmHg)
DBP predicted − true (mmHg)
mean of true and predicted SBP
mean of true and predicted DBP
bias -0.0
95% limits ±17
bias -0.9
95% limits ±18
bias +0.2
95% limits ±35
bias -8.0
95% limits ±43
bias -0.1
95% limits ±12
bias -0.7
95% limits ±12
bias +1.3
95% limits ±22
bias -5.5
95% limits ±26