
BP-bin weighting's cost and benefit, all five models
Cost: overall accuracy
weighted minus unweighted, positive = weighting costs accuracy
Overall SBP MAE change (mmHg)
2.5
2.0
1.5
1.0
0.5
0.0
-0.18
+1.92
+1.88
+0.74
+2.77
Random Forest
Gradient Boosting
1D-CNN
ResNet1D
Transformer
Benefit: high-BP coverage
weighted minus unweighted, negative = weighting reduces error
Above-170 mmHg SBP MAE change (mmHg)
0.0
−2.5
−5.0
−7.5
−10.0
−12.5
−15.0
−17.5
-4.05
-13.29
-16.73
-9.06
-16.41
Random Forest
Gradient Boosting
1D-CNN
ResNet1D
Transformer