

FilterCheck - 06 v1.1

Filter Test Report

Test Performed at : 2014-03-05, 12:02:16 PM

Serial No. : 3

Test Type : Bubble Point Test

Production Area : NA
Product Name : NA
Product Lot No. : NA

Filter Cat No. : NNXX1903XXXX103

Filter Lot No. : ND867672

Wetting Fluid : Water
Test Pressure : 14 psi
Results : 16.40 psi
Remarks : Test Passed

Mohit

(Test Operator)

2014-03-05