

## Filter Test Report

Test Performed at : 2014-03-04, 12:17:00 PM

Serial No.	: 1
Test Type	: Bubble Point Test
Production Area	: NA
Product Name	: NA
Product Lot No.	: NA
Filter Cat No.	: CNXX0902XXXX104
Filter Lot No.	: CB075752
Wetting Fluid	: Water
Test Pressure	: 32 psi
Results	: 36.02 psi
Remarks	: Test Passed

Mohit

(Test Operator)

2014-03-04