

Filter Test Report

Test Performed at : 2014-03-05, 02:44:16 PM

Serial No.	: 5
Test Type	: Bubble Point Test
Production Area	: NA
Product Name	: NA
Product Lot No.	: NA
Filter Cat No.	: LKSX5401EEXX101
Filter Lot No.	: LK1243-001
Wetting Fluid	: Water
Test Pressure	: 50 psi
Results	: 50.92 psi
Remarks	: Test Passed

Mohit

(Test Operator)

2014-03-05