

Filter Test Report

Test Performed on : 03-July-2014, 03:15:42 PM

Serial No.	: 131
Test Type	: Bubble Point Test
Production Area	: NA
Product Name	: NA
Product Lot No.	: NA
Filter Cat No.	: ITFX1001BBXX102
Filter Lot No.	: IT0742-044
Wetting Fluid	: 70% IPA
Pore Size	: 0.2 μ m
Filter Size	: 50 mm
Min. Bubble point value	: 22 psi
Test Value	: 26 psi
Test Result	: Test Passed

Sandeep
(Test Operator)
03-July-2014

(Test Operator)
(Test Operator)
03-July-2014