

## Filter Test Report

Test Performed at : 2014-03-04, 1:37:00 PM

Serial No.	: 2
Test Type	: Bubble Point Test
Production Area	: NA
Product Name	: NA
Product Lot No.	: NA
Filter Cat No.	: NNXX1901XXXX103
Filter Lot No.	: NA001012
Wetting Fluid	: Water
Test Pressure	: 50 psi
Results	: 54.03 psi
Remarks	: Test Passed

Mohit

(Test Operator)

2014-03-04