FICTIONAL DEMO DOCUMENT — Evidentra demo dossier. Not a real device, company, certificate or regulatory submission.

Device Master File — Summary

Document number: FMT-DMF-001

Revision: B

Document date: 12 March 2026

Device name: AuraPulse Fingertip Pulse Oximeter

Generic name: Pulse oximeter

Model number: AP-200A

Manufacturer: Fictional MedTech Pvt Ltd

Manufacturer address: Plot 12, Example Industrial Area, Pune 411000, India

Intended use: Spot-check, non-invasive measurement of functional oxygen saturation (SpO2) and pulse rate in adults.

Risk class (as determined by the manufacturer): Class B

Product Specification

SpO2 measurement range: 70% to 100%

SpO2 accuracy: ±2% (70% to 100%)

Pulse rate range: 30 to 250 bpm

Power supply: two AAA 1.5 V alkaline batteries

Display: OLED