FICTIONAL DEMO DOCUMENT — Evidentra demo dossier. Not a real device, company, certificate or regulatory submission.

Instructions for Use

Document number: FMT-IFU-001

Revision: A

Document date: 05 February 2026

AuraPulse Fingertip Pulse Oximeter

Model: AP-2000A

Manufactured by Fictional MedTech Pvt Ltd, Pune, India.

Intended use: Spot-check, non-invasive measurement of functional oxygen saturation (SpO2) and pulse rate in adults.

Specifications

SpO2 accuracy: ±2% (70% to 100%)

Pulse rate range: 30 to 250 bpm

Batteries: two AAA 1.5 V alkaline batteries

Warning: Do not use on neonates.