FICTIONAL DEMO DOCUMENT — Evidentra demo dossier. Not a real device, company, certificate or regulatory submission.

Risk Management Summary

Document number: FMT-RMS-001

Revision: A

Document date: 20 February 2026

Device: AuraPulse Fingertip Pulse Oximeter, model AP-200A

Risk analysis technique: preliminary hazard analysis.

Hazard 1: Inaccurate SpO2 reading due to low perfusion or patient movement.

Hazard 2: Skin irritation from prolonged finger contact.

Hazard 3: Battery depletion during measurement.

Risk control measures and residual risk evaluation: to be completed before design freeze.