FICTIONAL TEST DOSSIER — Evidentra. Example Biomedical Pvt Ltd and the OxiTrack OT-500 are not real. Not a regulatory submission.

Essential Principles Checklist

Essential Principles checklist for demonstrating conformity to the Essential Principles of Safety and Performance of the medical device OxiTrack OT-500.

Document number: EB-EPC-001 Revision: B Document date: 20 August 2026

Essential Principle	Applicable	Method of conformity	Document reference
General requirements — safety and performance	Yes	ISO 14971 risk management	EB-RMR-001
Chemical, physical and biological properties	Yes	ISO 10993-1, -5, -10	EB-BIO-003
Infection and microbial contamination	No	Non-sterile device for intact skin	—
Electrical safety and EMC	Yes	IEC 60601-1, IEC 60601-1-2	EB-TR-014
Devices with a measuring function	Yes	ISO 80601-2-61, desaturation study	EB-TR-014
Software	Yes	IEC 62304	EB-SW-002
Labelling and instructions for use	Yes	Chapter VI of MDR 2017	EB-LBL-001