FICTIONAL TEST DOSSIER — Evidentra. Example Biomedical Pvt Ltd and the OxiTrack OT-500 are not real. Not a regulatory submission.

Risk Management Report

Document number: EB-RMR-001 Revision: C Document date: 18 August 2026

Device: OxiTrack Fingertip Pulse Oximeter, model OT-500

Risk analysis technique: FMEA per ISO 14971:2019.

Hazard	Risk control	Residual risk
Inaccurate SpO2 reading (low perfusion, movement)	Signal-quality indicator; warning in IFU	Acceptable
Skin irritation from finger pad	Biocompatible silicone (ISO 10993)	Acceptable
Electric shock	Battery powered; IEC 60601-1 compliant design	Acceptable
Battery depletion during measurement	Low-battery indicator	Acceptable
Use on neonates	Contraindication in IFU	Acceptable

Overall residual risk evaluation: the overall residual risk is acceptable and the benefits outweigh the residual risks.