FICTIONAL TEST DOSSIER — Evidentra. Example Biomedical Pvt Ltd and the OxiTrack OT-500 are not real. Not a regulatory submission.

Label

Document number: EB-LBL-001 Revision: B Document date: 22 August 2026

OxiTrack Fingertip Pulse Oximeter

Model: OT-500

Manufactured by: Example Biomedical Pvt Ltd, Navi Mumbai, India

Batch: OT5-XXXX Mfg date / Expiry: see pack

Licence no.: to be issued

FICTIONAL TEST DOSSIER — Evidentra. Example Biomedical Pvt Ltd and the OxiTrack OT-500 are not real. Not a regulatory submission.

Instructions for Use

OxiTrack Fingertip Pulse Oximeter, model OT-500.

Intended use: Spot-check measurement of functional oxygen saturation (SpO2) and pulse rate in adults and paediatric patients in hospitals, clinics and the home.

Specifications

SpO2 measurement range: 70% to 100%

SpO2 accuracy: ±3% (70% to 100%)

Pulse rate range: 30 to 240 bpm

Batteries: two AAA 1.5 V alkaline batteries

Warnings

Do not use on neonates. Readings may be inaccurate with nail polish, low perfusion or movement.