Intended Use

The device is intended to secure an indwelling Foley urinary catheter to the patient's skin to reduce catheter movement and prevent accidental traction or dislodgement.

Indications for Use

The SureSet Foley Catheter Securement Device is indicated for use in adult patients with an indwelling Foley urinary catheter sized 14Fr to 18Fr. The device is intended for use by trained healthcare professionals, and by trained caregivers in non-clinical settings under the direction of a licensed clinician. The device is intended for use in hospital, ambulatory, long-term care, and home environments at ambient indoor temperature and humidity. Apply only to intact, clean, dry skin. Do not use on compromised skin, in environments with standing water exposure, or outside ambient indoor conditions.

Safety Considerations

Single use only. Do not reuse or reapply.
Discard if the package is damaged or the seal is broken.
Inspect the application site at least once per shift in clinical settings, and at least daily in home settings, for redness, swelling, drainage, or skin breakdown.

Key Usage Points

Prepare the skin. Cleanse with an antiseptic wipe and allow to dry completely before application.
Affix the device per the application steps in this document.
To access the catheter or tubing, place a finger under the clasp and lift to open the cover of the SureSet device.
Replace the device if the adhesive pad lifts, becomes soiled, or loses contact with the skin.
Remove any residue of the dressing pad adhesive with alcohol after removing the device.
Remove the device per the removal section.
Discard after use in accordance with local

Warnings, Precautions

Read and understand all warnings, precautions, and instructions for use prior to using the device.

⚠

Warnings

- Do not use the device where loss of adhesive could occur, such as non-adherent skin or confused patient.

- Use aseptic technique for application and removal of the device.

- In case of device loosening or malfunction, the device should be removed and replaced.

⚠

Precautions

- Read all package insert warnings and instructions prior to use.

- Use with caution in patient with known sensitivity to acrylate adhesives.

- The device and the surrounding area should be routinely inspected.

- The Foley securement device is to be used with 500 ml bags or smaller.

- Minimize catheter or tubing manipulation during application and removal of the device.

- Remove any oil and moisturizer from the targeted skin area before application.

- Respect infection control guidelines during application and removal of the device.

- Wash hands before use.

- This is a single use device. Do not reuse. If reused, the adhesive may not maintain its performance, causing the device to loosen

alera TES

MD **SureSet Foley**

Foley Catheter Securement and Stabilization Device

REF **SSF-30**

30 device(s) per box

UDI **00816022023985**

Alera TES
159 West Mason Avenue
Buckley, WA 98321
USA
+1(360) 829-0400
customerservice@aleramedtech.com
www.sureset-secure.com

Assembled in India
Mfg License Number:
MFG/MD/2023/001035

STERILE EO




IMPORTANT: Carefully read and understand all instructions, indications, warnings, precautions, and directions for use prior to using the SureSet Securement device. Failure to do so could result in compromised patient safety, patient complications, and/or insufficient treatment.

Visit our website to learn more


AT-100-L01-A Issue 2 Effective 09-Jul-2026

Device Overview



Instructions for Use


1




Clean the site where the device will be affixed.

Peel off the adhesive cover to reveal adhesive strip on underside

2




3




Place the device over the patient site ensuring it adheres securely.

Press and secure tubing into channel in the device (supported by tubing size -

4



5



Close the device cover. Make sure it closes securely.

Device Removal

To remove, open the cover and gently remove tubing from the channel. Peel the SureSet device gently off the skin. Dispose of the device according to hospital protocol or household waste guidelines, as appropriate to the setting.

R

