

BOEC ENGINEERING CONSULTANTS
SANAS Accredited Calibration Laboratory
No. CAL-2024-07
ISO/IEC 17025:2017 & ISO4037-3
Compliant

CERTIFICATE NO: 169e19f1
Date: 2024-11-14
Template: MCC15-07 v2.1
DIGITALLY SIGNED

1. Client Information

Client Name: Eskom Koeberg Nuclear Power Station
Address: Koeberg Rd, Melkbosstrand
Contact: Radiation Protection Manager

2. Instrument Under Test

Model: Thermo Fisher RadEye PRD-ER
Serial Number: PRD-ER-88471
Procedure: CP-02-08 / CP-03-07

3. Environmental Conditions (ISO4037-3 Clause 6.2)

Parameter	Measured Value	Required Limits	Status
Temperature	21.3 °C	18.0 °C – 24.0 °C	PASS
Relative Humidity	48.2 %RH	30.0 %RH – 60.0 %RH	PASS
Pressure	1012.4 hPa	860 hPa – 1060 hPa	PASS

4. Calibration Irradiation Results (Cs-137 Gamma Source)

Nominal Dose Rate	Measured Dose Rate	Response Factor	Uncertainty (k=2)	Compliance
0.500 mSv/h	0.512 mSv/h	1.024	± 4.5%	PASS
2.000 mSv/h	2.043 mSv/h	1.022	± 4.3%	PASS
10.00 mSv/h	10.18 mSv/h	1.018	± 4.2%	PASS

5. Metrological Traceability & RSA Digital Signature Seal

Metrological Traceability: Traceable to SI units held at NMISA (Cert: 2023-11-0445).
Technical Signatory: J. Van der Merwe (SANAS Auth: TM-042)
Approval Date: 2024-11-14

DIGITAL SIGNATURE VERIFIED
Signature: BOEC-RSA2048-SIG-AEBCD617E65D6643
SHA256: aebcd617e65d66439e95...
Status: RELEASED / PASSED