

BOEC ENGINEERING CONSULTANTS  
SANAS Accredited Calibration Laboratory  
No. CAL-2024-07  
ISO/IEC 17025:2017 & ISO4037-3  
Compliant

CERTIFICATE NO: 60203ac6  
Date: 2024-11-14  
Template: MCC15-07 v2.1  
DIGITALLY SIGNED

1. Client Information

Client Name: Eskom Koeberg Nuclear Power Station  
Address: Koeberg Rd, Melkbosstrand  
Contact: Radiation Protection Manager

2. Instrument Under Test

Model: Thermo Fisher RadEye PRD-ER  
Serial Number: PRD-ER-88471  
Procedure: CP-02-08 / CP-03-07

3. Environmental Conditions (ISO4037-3 Clause 6.2)

| Parameter         | Measured Value | Required Limits     | Status |
|-------------------|----------------|---------------------|--------|
| Temperature       | 21.3 °C        | 18.0 °C – 24.0 °C   | PASS   |
| Relative Humidity | 48.2 %RH       | 30.0 %RH – 60.0 %RH | PASS   |
| Pressure          | 1012.4 hPa     | 860 hPa – 1060 hPa  | PASS   |

4. Calibration Irradiation Results (Cs-137 Gamma Source)

| Nominal Dose Rate | Measured Dose Rate | Response Factor | Uncertainty (k=2) | Compliance |
|-------------------|--------------------|-----------------|-------------------|------------|
| 0.500 mSv/h       | 0.512 mSv/h        | 1.024           | ± 4.5%            | PASS       |
| 2.000 mSv/h       | 2.043 mSv/h        | 1.022           | ± 4.3%            | PASS       |
| 10.00 mSv/h       | 10.18 mSv/h        | 1.018           | ± 4.2%            | PASS       |

5. Metrological Traceability & RSA Digital Signature Seal

Metrological Traceability: Traceable to SI units held at NMISA (Cert: 2023-11-0445).  
Technical Signatory: J. Van der Merwe (SANAS Auth: TM-042)  
Approval Date: 2024-11-14

DIGITAL SIGNATURE VERIFIED  
Signature: BOEC-RSA2048-SIG-BC68417665D57597  
SHA256: bc68417665d575971f57...  
Status: RELEASED / PASSED