



ACCREDITED CALIBRATION LABORATORY

**Martin Calibration***"Your Partner in Quality"***Certificate of Calibration**

Certificate #: 4015405

**Calibration Performed By:**  
MARTIN CALIBRATION, LLC  
11965 12TH AVENUE SOUTH  
BURNSVILLE, MN 55337**For:**  
NORTECH SYSTEMS, INC.  
4050 NORRIS COURT NW  
BEMIDJI, MN 56601**Serial Number:** 13550375  
**Description:** THERMOMETER  
**Manufacturer:** FLUKE  
**Model:** 51 II  
**Cal Date:** 16 Feb 2026  
**Temp./RH:** 71 F / 33 %**Gage I.D.:** N1052**Performed By:** BRIAN ST GEORGE  
**Procedure:** SCP-160-1325  
**Due Date:** 16 Feb 2027Actual Measured Values Found to Be:  
**WITHIN NOMINAL TOLERANCE****Comments:****APPROVED**

02/23/2026

By: *John Bullington***Test Points**

| Description             | Set Point | Tolerance - | Tolerance + | As Found | Final  | Unit   | TUR  |
|-------------------------|-----------|-------------|-------------|----------|--------|--------|------|
| Display Test            | 1         | 1           | 1           | 1        | 1      | 1=Pass |      |
| Keypad Test             | 1         | 1           | 1           | 1        | 1      | 1=Pass |      |
| Temperature - K-Type T1 | 0.0       | -0.3        | 0.3         | 0.3      | 0.3    | °C     | 3.75 |
|                         | 23.0      | 22.7        | 23.3        | 22.7     | 22.7   | °C     | 3.75 |
|                         | -190.0    | -190.7      | -189.3      | -190.5   | -190.5 | °C     |      |
|                         | 990.0     | 989.2       | 990.8       | 989.9    | 989.9  | °C     |      |
|                         | 1200      | 1199        | 1201        | 1200     | 1200   | °C     | 1.72 |
|                         | 32.0      | 31.5        | 32.5        | 31.5     | 31.5   | °F     | 3.57 |
|                         | 73.4      | 72.9        | 73.9        | 72.9     | 72.9   | °F     | 3.57 |
|                         | -310.0    | -311.1      | -308.9      | -311.1   | -311.1 | °F     |      |
|                         | 1814      | 1813        | 1815        | 1814     | 1814   | °F     | 1.72 |
|                         | 2192      | 2190        | 2194        | 2192     | 2192   | °F     | 3.44 |

Decision Rule applied: Simple Acceptance; statements of compliance to specifications do not consider measurement uncertainty.

**Standards Used To Calibrate Equipment:**

| Serial Number | Gage ID    | Manufacturer | Model | Last Calibration | Due for Calibration |
|---------------|------------|--------------|-------|------------------|---------------------|
| 81568         | MI-160-524 | ECTRON CORP. | 1140A | 01 Apr 2025      | 30 Apr 2026         |





ACCREDITED CALIBRATION LABORATORY

## Certificate of Calibration

Certificate #: 4015405

Uncertainty in measurement: *(does not include thermocouple error)*

| Type J                  | Type T                  | Type K                  |
|-------------------------|-------------------------|-------------------------|
| (-210 to -180) °C: 0.14 | (-270 to -255) °C: 2.1  | (-270 to -255) °C: 2.5  |
| (-180 to -120) °C: 0.12 | (-255 to -240) °C: 0.56 | (-255 to -195) °C: 0.81 |
| (-120 to -50) °C: 0.09  | (-240 to -210) °C: 0.35 | (-195 to -115) °C: 0.14 |
| (-50 to 990) °C: 0.08   | (-210 to -150) °C: 0.21 | (-115 to -55) °C: 0.10  |
| (990 to 1200) °C: 0.08  | (-150 to -40) °C: 0.14  | (-55 to 1000) °C: 0.08  |
|                         | (-40 to 100) °C: 0.09   | (1000 to 1372) °C: 0.09 |
|                         | (100 to 400) °C: 0.08   |                         |

The above instrument was calibrated by standards traceable to the International System of Units (SI) through the National Institute of Standards & Technology (NIST) or other National Measurement Institute (NMI) per the guidelines specified in the latest revisions of ANSI/NCSL Z540-1 and ISO/IEC 17025. All measurement uncertainties calculated at a 95% Confidence Level ( $k=2$ ). All test points have a Test Uncertainty Ratio (TUR) of 4:1 or better unless otherwise noted. The results indicated on this certificate relate only to the items calibrated. This certificate shall not be reproduced, except in full, without the written approval of Martin Calibration. Martin Calibration's responsibility shall in no event, nor for any cause whatsoever, exceed the purchase price of this certificate.<sup>12</sup>

Quality Assurance Representative:

