

Characterization of nanoparticles by continuous contrast variation in SAXS

vorgelegt von
Master-Physiker
Raul Garcia Diez
geboren in Barcelona, Spanien

Von der Fakultät II - Mathematik und Naturwissenschaften
der Technischen Universität Berlin
zur Erlangung des akademischen Grades
Doktor der Naturwissenschaften
Dr. rer. nat.

genehmigte Dissertation

Promotionsausschuss:

Vorsitzender: Prof. Dr. XXX

Gutachter: Prof. Dr. Stefan Eisebitt

Gutachterin: Prof. Dr. Simone Raoux

Gutachter: Prof. Dr. Mathias Richter

Tag der wissenschaftlichen Aussprache:

Berlin 2016

1

Introduction

In 1966, Richard Fleischer directed *Fantastic Voyage*, a film about the voyage of a miniaturized submarine used to cruise along human blood vessels and repair the damaged caused to the scientist's brain by a blood clot. The idea of treating from the inside damaged cells or organs fuelled the imagination of the next generation scientists and shaped the incipient field of nanomedicine. Less than 30 years later, science fiction became science fact and Doxil was approved by the US Food and Drug Administration in 1995 as the first nanodrug commercially available (Barenholz, 2012). Although 20 years after this milestone nano-submarines are still a long way off, nanomedicine is a well-established research field and dozens of products are under clinical trials or have been approved by the relevant health agencies.

The precursor to the nanomedicine breakthrough can be found in the tremendous progress in nanoparticles research observed in the 60s and 70s of the last century. Nanoparticles (NPs) are objects with *one or more external dimensions in the size range from 1 nm to 100 nm* (European Commission Recommendation for nanomaterial (2011/696/EU)) and have a preeminent position in the continuously growing world of nanotechnology, employed as paints or cosmetic products (Guterres *et al.*, 2007). Besides, the application of NPs in the emerging field of nanomedicine opens up exciting prospects (Nie *et al.*, 2007; Sahoo & Labhasetwar, 2003; Wickline & Lanza, 2003; Zhou *et al.*, 2014; Rosen & Abribat, 2005), especially considering their possibilities as platforms for drug-delivery (Wang *et al.*, 2012) or encapsulating imaging agents (Tao *et al.*, 2011).

The development of NPs is currently focused towards tailoring nanodrug carriers with flexible surface functionalisations and controlled morphologies (Euliss *et al.*, 2006; Yang *et al.*, 2005). The morphology of NPs is typically specified by parameters like size, shape, density or chemical composition of the particle, which are fundamental and defining aspects of the particle functions and determine their applicability in real-world medical applications (Vittaz *et al.*, 1996). In this regard, the size of NPs is one of the most crucial physicochemical properties of nanodrugs, because it delimits whether they can intrude into the biological cells or the targeted tumor sites. An accurate and reliable description of the morphological traits of the NPs is therefore of vital importance for their favorable translation into successful nanomaterials.




Figure 1.1 | Relevant properties of the nanoparticles.

The term *nanometrology* refers to the science of accurate and correct measurement of relevant properties at the nanometre range. A central concept in metrology is *traceability*, which refers to the ability of relating the measured value i.e. measurand to a base unit definition of the International System of Units (SI system) by an unbroken chain of comparisons with known uncertainties. This enables an objective comparison of the results obtained by different methods based on a consistent uncertainty budget associated to the measurand. The fundamental research in the field of metrology in Germany is addressed by its national metrology institute, the Physikalisch-Technische Bundesanstalt (PTB). Founded in 1887, the PTB is devoted among other metrological activities to the determination of fundamental and natural constants or the technology transfer for the industry.

In the nanoscale level, PTB is involved in the development of the dimensional nanometrology field, which studies the measurement of the physical size or distances of a given nanomaterial and traces it back to the *nm* unit. There are several available techniques which are suitable for the sizing of NPs, though not all provide a traceable measurement. A prime example is dynamic light scattering (DLS), the most widely used tool in nanomedicine (Murphy, 1997; Hallett *et al.*, 1991; Egelhaaf *et al.*, 1996; Takahashi *et al.*, 2008; Jans *et al.*, 2009; Hoo *et al.*, 2008). DLS is well-established and has indisputable advantages in the size characterization of the NPs, e.g. easy-to-use instrumentations, fast and low-cost operation, but it is not capable of a traceable size determination as there is no general relationship between the measured hydrodynamic diameter and the physical size of the NPs (Meli *et al.*, 2012).

Other ensemble techniques extensively used are differential centrifugal sedimentation (DCS) (Fielding *et al.*, 2012) and Nanoparticle Tracking Analysis (NTA), both capable of measuring the NPs in suspension, i.e. in their native medium. While DCS is based on the sedimentation of NPs through a density gradient, NTA is a single-particle counting method that relates the Brownian movement of the particles with the measured laser

light scattering. The particle size distribution obtained with DCS is traced down to a calibration standard of known size and density, resulting in a non-traceable measurement. Similarly to DLS, the NTA measurand derives from the hydrodynamic properties of the NPs (REF-Towards traceable size determination of extracellular vesicles).

Microscopic tools are also frequently used for structural investigations (Joensson *et al.*, 1991; Silverstein *et al.*, 1989) and proved to be useful techniques for solid NPs due to their SI traceability achieved by coupling the measurement table with a laser interferometer (Meli *et al.*, 2012). Nevertheless, techniques such as transmission scanning electron microscopy (TSEM), transmission X-ray microscopy (TXM) or atomic force microscopy (AFM) are not ensemble averaged and the statistical accuracy of non-ensemble methods is usually not sufficient. Besides, the removal of the original suspending medium can be considered another drawback, as well as the possible distortion of the particle's morphology during the drying process, though it can be partially overcome by cryo-TEM (Li *et al.*, 1998).




Figure 1.2 | Some available sizing techniques for nanoparticles and their available size range.

The nanoparticles envisioned for medical use tend to be found in the soft matter regime and thus the characterization tools must be carefully chosen considering the measurement limitations. For example, biodegradable NPs, e.g. polymeric colloids, are finding many medical applications, especially as drug-carriers (Kattan *et al.*, 1992; Vicent & Duncan, 2006) and are starting to undergo clinical trials (Patel *et al.*, 2012; Beija *et al.*, 2012; Cabral & Kataoka, 2014). However, the size determination of polymeric NPs with a well-known technique like AFM is rather challenging (REF-Particle deformation induced by AFM tapping under different setpoint voltages) and suggests alternative approaches.

Liposomes are spheric vesicles composed of a closed phospholipid bilayer membrane capable of encapsulating hydrophilic compounds. The importance of lipid vesicles in the progress of nanomedicine is indisputable, as the first approved nanodrug is a liposomal formulation of doxorubicin, Doxil. Nowadays liposomes continue to be a widespread instrument for drug delivery (Pérez-Herrero & Fernández-Medarde, 2015), but their complicate internal structure requires typically more than a single a characterization tool (Khorasani *et al.*, 2014). Likewise, relevant biological structures in nanomedicine possess heterogenous morphologies rather difficult to detect with imaging techniques.